

Name: Worlow
Next test due:

TSI MODEL 8015 PORTACOUNT FIT-TESTING SOFTWARE
Version 2.01 S/N 927
VISTA POLYMERS

FIT TEST REPORT

Test date: 08-09-1994
Test time: 7:34:24 am
Test subject last name: Worlow
Test subject first name: Charles
ID number: 426941635
Operator name: H Pruitt
Next fit test due:
Test agent: Ambient particles
Test device: TSI MODEL 8010 PORTACOUNT

RESPIRATOR ID

Size: Medium
Model: Ultravue
Manufacturer: Facepiece
Approval number: TC-19C-113

Additional notes:
Additional notes:

TEST DATA

Pass/Fail level: 100
Test cycles.....: 1

Ex.	Ambient (Part/cc)	Mask (Part/cc)	Fit Factor	Pass/Fail
NB	6110	1.180	5180	PASS
DB	6210	0.550	11300	PASS
SS	6355	1.170	5430	PASS
UD	6300	0.650	9700	PASS
R	6185	8.130	761	PASS
NB	6230	2.510	2480	PASS

OVERALL FIT FACTOR = 2625 PASS
(Calculated per ANSI Z88.2-1980)

Operator Helene Pruitt Date 8-9-94
Test Subject Charles Worlow Date 8-9-94