

Name: Frye  
Next test due:

TSI MODEL 8015 PORTACOUNT FIT-TESTING SOFTWARE  
Version 2.01 S/N 927  
VISTA POLYMERS

FIT TEST REPORT

-----  
Test date: 08-02-1994  
Test time: 2:02:07 pm  
Test subject last name: Frye  
Test subject first name: Larry  
ID number: 416741049  
Operator name: S. Carter  
Next fit test due:  
Test agent: Ambient particles  
Test device: TSI MODEL 8010 PORTACOUNT

RESPIRATOR ID  
Size: Small  
Model: Ultravue Facepiece  
Manufacturer: MSA  
Approval number: TC-19C-113

Additional notes:  
Additional notes:

TEST DATA

-----  
Pass/Fail level: 100  
Test cycles.....: 1

| Ex. | Ambient<br>(Part/cc) | Mask<br>(Part/cc) | Fit Factor | Pass/Fail |
|-----|----------------------|-------------------|------------|-----------|
| NB  | 2455                 | 32.000            | 76.8       | FAIL      |
| DB  | 2355                 | 18.500            | 127        | PASS      |
| SS  | 3105                 | 3.710             | 837        | PASS      |
| UD  | 3915                 | 1.220             | 3210       | PASS      |
| R   | 5050                 | 28.800            | 175        | PASS      |
| NB  | 5285                 | 2.430             | 2170       | PASS      |

OVERALL FIT FACTOR = 210 PASS  
(Calculated per ANSI Z88.2-1980)

Operator Shelby D. Carter Date 8-2-94  
Test Subject Larry Frye Date 8-2-94