

AR226-1678

Sample Id	units	GC ECD (Historic SOP)	GC MS (Proposed SOP)
GALLERY	ug/L	3.99	18.3
RANNEY	ug/L	5.48	30.9
WEST WELFIELD #1	ug/L	1.58	4.15
GALLERY	RPR PFNA	18	97
RANNEY	RPR PFNA	16	101
WEST WELFIELD #1	RPR PFNA	18	80
BLANK	RPR PFNA	12	110
LCS	RPR PFNA	9	96
LCS D	RPR PFNA	NA	91
MS	RPR PFNA	16	NA
MSD	RPR PFNA	19	NA

notes: RPR PFNA = relative percent recovery (%) for the surrogate standard perfluorononanoic acid which is spiked into each sample and associated quality control sample during sample preparation for analysis

LCS/LCSD = laboratory control spike/laboratory control spike duplicate

MS/MSD = matrix spike/matrix spike duplicate

NA = sample not included in batch analysis

prepared for John Bowman on 5/17/01

133

000004000

BEST COPY AVAILABLE

000005