

Name: Spann
Next test due:

TSI MODEL 8015 PORTACOUNT FIT-TESTING SOFTWARE
Version 2.01 S/N 927
VISTA POLYMERS

FIT TEST REPORT

Test date: 08-12-1994
Test time: 11:02:06 am
Test subject last name: Spann
Test subject first name: Doug
ID number: 418663073
Operator name: S. Carter
Next fit test due:
Test agent: Ambient particles
Test device: TSI MODEL 8010 PORTACOUNT

RESPIRATOR ID

Size: Medium
Model: Ultravue Facepiece
Manufacturer: MSA
Approval number: TC-19C-113

Additional notes:
Additional notes:

TEST DATA

Pass/Fail level: 100
Test cycles.....: 1

Ex.	Ambient (Part/cc)	Mask (Part/cc)	Fit Factor	Pass/Fail
NB	26650	0.290	91900	PASS
DB	26850	0.760	35300	PASS
SS	26800	0.080	335000	PASS
UD	26900	0.050	539000	PASS
R	26650	5.240	5100	PASS
NB	26450	0.160	165000	PASS

OVERALL FIT FACTOR = 24371 PASS
(Calculated per ANSI Z88.2-1980)

Operator Shelby D. Carter Date 8-12-94
Test Subject Douglas M. Spann Date _____