

Name: Gann
Next test due:

TSI MODEL 8015 PORTACOUNT FIT-TESTING SOFTWARE
Version 2.01 S/N 927
VISTA POLYMERS

FIT TEST REPORT

Test date: 08-08-1994
Test time: 7:48:01 am
Test subject last name: Gann
Test subject first name: Joe
ID number: 428524358
Operator name: H Pruitt
Next fit test due:
Test agent: Ambient particles
Test device: TSI MODEL 8010 PORTACOUNT

RESPIRATOR ID

Size: Large
Model: Ultravue
Manufacturer: Facepiece
Approval number:

Additional notes:
Additional notes:

TEST DATA

Pass/Fail level: 100
Test cycles.....: 1

Ex.	Ambient (Part/cc)	Mask (Part/cc)	Fit Factor	Pass/Fail
NB	7975	0.460	17300	PASS
DB	7965	1.000	7970	PASS
SS	7860	0.190	41400	PASS
UD	7975	0.970	8220	PASS
R	8130	2.620	3100	PASS
NB	8175	0.150	54500	PASS

OVERALL FIT FACTOR = 8955 PASS
(Calculated per ANSI Z88.2-1980)

Operator Helen Pruitt Date 8-8-94

Test Subject Joe a Gann Date 8-8-94