

Name: EVANS
Next test due:

TSI MODEL 8015 PORTACOUNT FIT-TESTING SOFTWARE
Version 2.01 S/N 927
VISTA POLYMERS

FIT TEST REPORT

Test date: 10-01-1994
Test time: 8:43:59 pm
Test subject last name: EVANS
Test subject first name: REGINALD
ID number: 425196023
Operator name: STEWART
Next fit test due:
Test agent: Ambient particles
Test device: TSI MODEL 8010 PORTACOUNT

RESPIRATOR ID

Size: MEDIUM
Model: ULTRAVUE FACEPIECE
Manufacturer: 3M
Approval number: TC-21C-488

Additional notes:
Additional notes:


TEST DATA

Pass/Fail level: 100
Test cycles.....: 1

Ex.	Ambient (Part/cc)	Mask (Part/cc)	Fit Factor	Pass/Fail
NB	7075	0.790	8960	PASS
DB	8265	2.080	3970	PASS
SS	8795	0.730	12000	PASS
UD	8805	0.730	12000	PASS
R	8425	6.700	1250	PASS
NB	7810	0.400	19500	PASS

OVERALL FIT FACTOR = 4343 PASS
(Calculated per ANSI Z88.2-1980)

Operator



Date

10-1-94

Test Subject



Date

10/1/94