

"TEST START DATE: 8-9-89"
 "TEST START TIME: ----"
 "TEST LOCATION: CALENDER"
 "EMPLOYEE NAME: RHONDA COLEMAN"
 "EMPLOYEE NUMBER: 150-40-5000"
 "DEPARTMENT: CALENDER PRODUCTION"
 "COMMENT FIELD 1: OPERATOR, BANBURY #2"
 "COMMENT FIELD 2: #'S 1, 2, 3 RUNNING"
 "NUMERIC CODES: ----"

METROSONICS db-308 SN 2528 V2.3
 REPORT PRINTED 08/09/89 @ 15:45:01

DOUBLING RATE: 5dB FILTER: A WGH
 DOSE CRITERION: 90dB RESPONSE: SLOW
 PRE-CALIBRATION TIME: 8/09/89 @ 8:09:51
 PRE-CALIBRATION RANGE: 43.2dB TO 139.2dB

CALIBRATOR TYPE & SERIAL #: -----

CALIBRATOR CALIBRATION DATE: -----

TEST BEGAN 8/09/89 @ 8:20:19
 TEST LENGTH: 0DAYS 07:10:00

Lav = 88.3dB
 SEL = N/A
 Lmax = 121.7dB ON 8/09/89 @ 11:24:06
 Lpk = OVER ON 8/09/89 @ 10:51:21
 TIME OVER 115dB 0D 0:00:03.87

8 HR DOSE (80dB CUTOFF) = 70.23%
 8 HR PROJ. DOSE (80dB CUTOFF) = 78.39%
 8 HR DOSE (90dB CUTOFF) = 33.13%
 8 HR PROJ. DOSE (90dB CUTOFF) = 36.98%

Sound Level Meter Calibration

SLM Mfg.

SN 2528

Pre-Cal					
Calibrator SN 12185 Gen Rad 1562-A					
Location <i>CALENDER</i>					
Results					
	125	250	500	1000	2000
dBA				114.0	
dBC					
Initials <i>MTS</i>		Date/Time 8/9/89 8:05			

Post-Cal					
Calibrator SN 12185 Gen Rad 1562-A					
Location <i>CALENDER</i>					
Results					
	125	250	500	1000	2000
dBA				114.1	
dBC					
Initials <i>MTS</i>		Date/Time 3:48pm 8-9-89			

OCC 020386