5

MONSANTO CHEMICA CO ANNISTON

MONSANTO STL 4-2 EK
FROM PLANT B
GERBER
WE FIND BOTH THE VISCOSIMETER AND MEGOHM SETUP SANXXX SATISFACTORY
THEY ARE BOTH CONSTANTLY IN USE THE VISCOSIMETER IS SPECIFIED
FOR G E TESTS IT MIGHT BE RECOMMENDED THAT A POWER PACK SOURCE
BE USED INSTEAD OF BATTERIES ON THE MEGOHM METER
WIEGAND

9

END
ENDV

DSW 331098