



Right Science. Right People.

UPSTATE NEW YORK:

7500 West Henrietta Road
Rush, NY 14543
p: 800.836.4850 • 585.533.1672
f: 585.533.1796

WESTERN PENNSYLVANIA:

2205 East 33rd Street
Erie, PA 16510
p: 800.836.4850 • 814.899.4500
f: 814.899.7089

USP <87> Compliance Certificate

Sponsor: Garlock Sealing Technologies
1666 Division Street
Palmyra, NY 14522-9350

Attn: Michael McNally

Test No: T16-3344

Test Material: GYLON Style 3504 (Branded)

This test material has met the requirements of USP <87> Biological Reactivity Tests, In Vitro; Elution Test, in accordance with USP 39 – NF 34 through Second Supplement, Official December 1, 2016 to April 30, 2017.


Study Director

1-9-17
Date



bioscience
Right Science. Right People.

UPSTATE NEW YORK:

7500 West Henrietta Road
Rush, NY 14543
p: 800.836.4850 • 585.533.1672
f: 585.533.1796

WESTERN PENNSYLVANIA:

2205 East 33rd Street
Erie, PA 16510
p: 800.836.4850 • 814.899.4500
f: 814.899.7089

USP Class VI Plastic Designation

Sponsor: Garlock Sealing Technologies
1666 Division Street
Palmyra, NY 14522-9350

Attention: Michael McNally

Test Nos.: T16-3345, 3346, 3347

Test Material: Gylon Style 3504 (Branded)

This test material meets the requirements of USP 39 – NF 34, General Chapter
<88> Biological Reactivity Tests, In Vivo for Class VI designation at a temperature
extraction of 121°C for 1 hour.

Tonia Schaubroek

Study Director

1-24-17

Date