

Name: Whatley
Next test due:

TSI MODEL 8015 PORTACOUNT FIT-TESTING SOFTWARE
Version 2.01 S/N 927
VISTA POLYMERS

FIT TEST REPORT

Test date: 08-09-1994
Test time: 11:10:57 am
Test subject last name: Whatley
Test subject first name: Donald
ID number: 451800091
Operator name: H Pruitt
Next fit test due:
Test agent: Ambient particles
Test device: TSI MODEL 8010 PORTACOUNT

RESPIRATOR ID

Size: Medium
Model: Ultravue
Manufacturer: Facepiece
Approval number: TC-19C-113

Additional notes:
Additional notes:

TEST DATA

Pass/Fail level: 100
Test cycles.....: 1

Ex.	Ambient (Part/cc)	Mask (Part/cc)	Fit Factor	Pass/Fail
NB	8580	0.530	16100	PASS
DB	9205	0.210	43800	PASS
SS	9660	0.220	43900	PASS
UD	9790	0.360	27200	PASS
R	9800	3.810	2570	PASS
NB	9660	2.020	4780	PASS

OVERALL FIT FACTOR = 8078 PASS
(Calculated per ANSI Z88.2-1980)

Operator Helen Pruitt Date 8-9-94
Test Subject Donald R. Whatley Date 8-9-94