

Name: COGSDELL
Next test due:

TSI MODEL 8015 PORTACOUNT FIT-TESTING SOFTWARE
Version 2.01 S/N 927
VISTA POLYMERS

FIT TEST REPORT

Test date: 09-30-1994
Test time: 9:33:24 pm
Test subject last name: COGSDELL
Test subject first name: PAUL
ID number: 587364460
Operator name: STEWART
Next fit test due:
Test agent: Ambient particles
Test device: TSI MODEL 8010 PORTACOUNT

RESPIRATOR ID
Size: MEDIUM
Model: ULTRAVUE FACEPIECE
Manufacturer: 3M
Approval number: TC-31C-488

Additional notes:
Additional notes:

TEST DATA

Pass/Fail level: 100
Test cycles.....: 1

Ex.	Ambient (Part/cc)	Mask (Part/cc)	Fit Factor	Pass/Fail
NB	10395	2.220	4690	PASS
DB	11150	1.540	7270	PASS
SS	11050	1.440	7710	PASS
UD	11050	0.550	20100	PASS
R	10950	6.220	1760	PASS
NB	10900	0.720	15200	PASS

OVERALL FIT FACTOR = 5154 PASS
(Calculated per ANSI Z88.2-1980)

Operator Arthur J. Stewart Date 9-30-94
Test Subject Paul Cogsdell Date 9-30-94