

Name: Gilreath
Next test due:

TSI MODEL 8015 PORTACOUNT FIT-TESTING SOFTWARE
Version 2.01 S/N 927
VISTA POLYMERS

FIT TEST REPORT

Test date: 08-09-1994
Test time: 9:20:56 am
Test subject last name: Gilreath
Test subject first name: Christopher
ID number: 587378334
Operator name: H Pruitt
Next fit test due:
Test agent: Ambient particles
Test device: TSI MODEL 8010 PORTACOUNT

RESPIRATOR ID

Size: Medium
Model: Ultravue
Manufacturer: Facepiece
Approval number: TC-19C-113

Additional notes:
Additional notes:

TEST DATA

Pass/Fail level: 100
Test cycles.....: 1

Ex.	Ambient (Part/cc)	Mask (Part/cc)	Fit Factor	Pass/Fail
NB	7220	0.320	22500	PASS
DB	7245	0.150	48300	PASS
SS	7260	0.660	11000	PASS
UD	7210	0.700	10300	PASS
R	7180	0.920	7800	PASS
NB	7095	0.430	16500	PASS

OVERALL FIT FACTOR = 13576 PASS
(Calculated per ANSI Z88.2-1980)

Operator Helen Pruitt Date 8-9-94

Test Subject Chris H. Pruitt Date 8-9-94