

Name: Carroll
Next test due:

TSI MODEL 8015 PORTACOUNT FIT-TESTING SOFTWARE
Version 2.01 S/N 927
VISTA POLYMERS

FIT TEST REPORT

Test date: 08-12-1994
Test time: 7:52:21 am
Test subject last name: Carroll
Test subject first name: Steve
ID number: 426354963
Operator name: S. Carter
Next fit test due:
Test agent: Ambient particles
Test device: TSI MODEL 8010 PORTACOUNT

RESPIRATOR ID

Size: Medium
Model: Ultravue Facepiece
Manufacturer: MSA
Approval number: TC-19C-113

Additional notes:
Additional notes:

TEST DATA

Pass/Fail level: 100
Test cycles.....: 1

Ex.	Ambient (Part/cc)	Mask (Part/cc)	Fit Factor	Pass/Fail
NB	40050	0.120	334000	PASS
DB	41300	0.300	137000	PASS
SS	41300	0.290	142000	PASS
UD	41050	0.360	114000	PASS
R	41150	4.780	8610	PASS
NB	42750	0.110	389000	PASS

OVERALL FIT FACTOR = 41430 PASS
(Calculated per ANSI Z88.2-1980)

Operator Shelby D. Carter Date 8-12-94
Test Subject WB Carroll Date 8/12/94