

Name: Minor
Next test due:

TSI MODEL 8015 PORTACOUNT FIT-TESTING SOFTWARE
Version 2.01 S/N 927
VISTA POLYMERS

FIT TEST REPORT

Test date: 08-12-1994
Test time: 8:26:19 am
Test subject last name: Minor
Test subject first name: Jerry
ID number: 416505295
Operator name: S. Carter
Next fit test due:
Test agent: Ambient particles
Test device: TSI MODEL 8010 PORTACOUNT

RESPIRATOR ID
Size: Medium
Model: Ultravue Facepiece
Manufacturer: MSA
Approval number: TC-19C-113

Additional notes:
Additional notes:

TEST DATA

Pass/Fail level: 100
Test cycles.....: 1

Ex.	Ambient (Part/cc)	Mask (Part/cc)	Fit Factor	Pass/Fail
NB	28000	0.940	29800	PASS
DB	26450	1.090	24300	PASS
SS	24700	0.750	33000	PASS
UD	22800	0.290	78800	PASS
R	21300	19.500	1090	PASS
NB	19750	1.430	13800	PASS

OVERALL FIT FACTOR = 5417 PASS
(Calculated per ANSI Z88.2-1980)

Operator Shelby D. Carter Date 8-12-94

Test Subject Jerry Minor Date 8-12-94