

MEASUREMENT OF  
AIRBORNE ASBESTOS FIBER  
BY THE  
MEMBRANE FILTER METHOD



ASBESTOS TEXTILE INSTITUTE

(FOUNDED 1944)

HER 0001573

©Asbestos Textile Institute 1971

Printed in U.S.A.

Price: \$1.00 per copy

|                                     |
|-------------------------------------|
| SCOPE .....                         |
| DEFINITION .....                    |
| OUTLINE OF METHOD .....             |
| APPARATUS AND MATERIALS .....       |
| CLEANLINESS .....                   |
| PREPARATION OF MOUNTING SOLUTION    |
| AEROSOL FIELD MONITOR AND FILTERS   |
| CALIBRATION OF EQUIPMENT .....      |
| SAMPLING .....                      |
| COUNTING .....                      |
| CALCULATION OF RESULTS .....        |
| INDUSTRIAL HYGIENE WORK SHEET ..... |
| BIBLIOGRAPHY .....                  |

## CONTENTS


|                                         | Page |
|-----------------------------------------|------|
| SCOPE .....                             | 5    |
| DEFINITION .....                        | 5    |
| OUTLINE OF METHOD .....                 | 5    |
| APPARATUS AND MATERIALS .....           | 5    |
| CLEANLINESS .....                       | 7    |
| PREPARATION OF MOUNTING SOLUTION .....  | 7    |
| AEROSOL FIELD MONITOR AND FILTERS ..... | 7    |
| CALIBRATION OF EQUIPMENT .....          | 7    |
| SAMPLING .....                          | 10   |
| COUNTING .....                          | 11   |
| CALCULATION OF RESULTS .....            | 14   |
| INDUSTRIAL HYGIENE WORK SHEET .....     | 15   |
| BIBLIOGRAPHY .....                      | 16   |

tile Institute 1971

ed in U.S.A.

\$1.00 per copy

SAMPLING EQUIPMENT  
USED IN THE  
MEMBRANE FILTER METHOD



4

HER 0001576

## 1. SCOPE:

- 1.1 This method covers a procedure for fiber concentration in air. The procedure includes the determination of concentration using phase contrast microscopy.

This method is also useful for fibers less than 5 microns, the minimum length by the inherent limitations of ordinary contrast optics. Further, this method is used only in areas where procedures for distinguishing between asbestos fibers and non-asbestos fibers are required. All fibers detected by this procedure are reported as such, unless warranted by other assumptions.

This method can be used to assess fiber concentrations to standards as set forth in provincial statutory requirements for Asbestos.

## 2. DEFINITION:

- 2.1 A fiber is any particle having a length greater than 5 microns.

## 3. OUTLINE OF METHOD:

- 3.1 Airborne asbestos fibers are collected from air through a membrane filter by which fibers are deposited on the filter.

A representative section of the filter is mounted on a microscope slide, and examined (or 400x) using phase contrast optics. The number of fibers observed in a given area under this figure is used to calculate the number of fibers per cubic meter (cc.) of air sampled.

Fiber concentrations are categorized in the following manner; total number of fibers, fibers greater than 5 microns, and fibers greater than 10 microns per cubic meter.

## 4. APPARATUS AND MATERIALS:

- 4.1 Sample Collector:  
The samples shall be collected on a filter.

5

## 1. SCOPE:

- 1.1 This method covers a procedure for the determination of asbestos fiber concentration in air. The procedure is designed specifically for the determination of concentrations of fibers 5 microns or longer using phase contrast microscopy.

This method is also useful for the determination of fibers shorter than 5 microns, the minimum length detectable being limited only by the inherent limitations of ordinary light microscopy using phase contrast optics. Further, this method is meant to be used for fiber determinations only in areas where airborne fibers, if present, are predominantly or entirely asbestos fibers; it does not specify procedures for distinguishing between asbestos and other types of fibers. All fibers detected by this procedure are assumed to be asbestos fibers and are reported as such, unless independent evidence is at hand to warrant other assumptions.

This method can be used to assess conformance of asbestos fiber concentrations to standards as set forth by Federal, State, or Provincial statutory requirements for Asbestos Threshold Limit Value.

## 2. DEFINITION:

- 2.1 A fiber is any particle having a length to diameter ratio of 3:1 or greater.

## 3. OUTLINE OF METHOD:

- 3.1 Airborne asbestos fibers are collected by drawing a specified volume of air through a membrane filter by use of a suction pump. The fibers are deposited on the filter.


A representative section of the filter is removed, properly mounted on a microscope slide, and examined at a magnification of 430x (or 400x) using phase contrast optics. The number and length range of fibers observed in a given area under the microscope is noted and this figure is used to calculate the number of fibers per cubic centimeter (cc.) of air sampled.

Fiber concentrations are categorized with respect to length in the following manner; total number of fibers seen, greater than 5 microns, and greater than 10 microns per cubic centimeter of air.

## 4. APPARATUS AND MATERIALS:

- 4.1 Sample Collector:

The samples shall be collected on a filter having a pore size of  $0.8 \pm$



0.05 micrometers (microns), supported on a thick pad to control air distribution and mounted in a segmented enclosure (filter holder, field monitor) to prevent contamination of the filter prior to exposure.(1)

#### 4.2 Cellulose Bands:

The segments of the filter holder (field monitor) containing the filter and pad shall remain sealed against contamination until ready for examination.(2)

#### 4.3 Tubing:

Rubber or plastic tubing of suitable length is connected to the inlet side of the pump and to the filter holder assembly. One quarter inch inside diameter, non-collapsible tubing is recommended.

#### 4.4 The suction pump shall be capable of providing a constant air flow rate of 2 liters per minute for the duration of a test when connected to the filter assembly through a required length of tubing. It is desirable, although not necessary, that the pump have an explosion-proof motor.(3)

#### 4.5 Microscope and Accessories:

A binocular microscope fitted with phase contrast optics should be used for counting the asbestos fibers. Any reticle, such as Porton or Patterson Globe and Circle, which projects a constant counting area and has provision for sizing, is suitable. This procedure is written using the Bausch and Lomb phase-equipped microscope and the Porton reticle. All set up procedures correspond with this equipment. However, any good microscope with phase contrast accessories is acceptable.

The reticle is mounted in the non-adjustable eyepiece and a stage micrometer is used to calibrate the dimensions of the counting area.

Microscope slides size 3 x 1 inch and cover glasses No. 1½, 18mm. must be used.

(1) A suitable filter assembly is a Type AA, 37mm., 0.8 micron cellulose ester filter, pad and plastic "field monitor" — obtainable from the Millipore Corporation, Bedford, Mass. 01730.

(2) Cellulose bands for this purpose may be obtained from the W. H. Jelly and Co., Inc., Franklin Park, Illinois 60131.

(3) Suitable pumps may be obtained from: a) Mine Safety Appliance Co., Pittsburgh, Penna. 15208; b) Willson Products, Div., Reading, Penna. 19603; c) Unico Environmental Instruments Inc., Fall River, Mass. 02720.

#### 4.6 Miscellaneous:

A 1:1 mixture of dimethyl phthalate for preparing the sample mounting medium

Tweezers, scalpel, lens tissue, and cleaning cleanliness of the microscope and

### 5. CLEANLINESS:

5.1 All equipment used in fiber counting only for this purpose. Fiber counting part of the laboratory; and, if possible, microscope and accessories should be cleaned according to the procedures for care of the microscope and manufacturer's recommendations.

### 6. PREPARATION OF MOUNTING SOLUTION

6.1 The mounting solution used in this procedure is dimethyl phthalate and diethyl oxalate when the sample is to be counted with a microscope. If samples are to stand for longer periods, a fixative should be used. This solution may be prepared by dissolving 1 gram of membrane filter material per 100 ml of the dissolved material is to provide a viscosity as possible without being too viscous. A viscous solution delays the migration of the sample to the center of the mount. The sample may show an apparent loss in concentration for application.

### 7. AEROSOL FIELD MONITOR AND FILTER

7.1 The monitors may be reused after cleaning thoroughly in warm detergent water and rinsing.

### 8. CALIBRATION OF EQUIPMENT:

#### 8.1 Pump:

The frequency of calibration is dependent on the handling of the instrument. The calibration is carried out with a filter in place. It is recommended that an intermediate standard be used in the calibration.

8.1.1 Suction pumps equipped with a flowmeter calibrated by a primary standard procedure is shown in Fig. 1. A direct measure of air volume is

ns), supported on a thick pad to control air  
ed in a segmented enclosure (filter holder,  
at contamination of the filter prior to ex-

holder (field monitor) containing the filter  
led against contamination until ready for

of suitable length is connected to the inlet  
the filter holder assembly. One quarter inch  
psable tubing is recommended.

be capable of providing a constant air flow  
e for the duration of a test when connected  
rough a required length of tubing. It is  
ecessary, that the pump have an explosion-

es:  
itted with phase contrast optics should be  
estos fibers. Any reticle, such as Porton or  
e, which projects a constant counting area  
ing, is suitable. This procedure is written  
omb phase-equipped microscope and the  
rocedures correspond with this equipment.  
oscope with phase contrast accessories is

the non-adjustable eyepiece and a stage  
rate dimensions of the counting area.

1 inch and cover glasses No. 1½, 18mm.

Type AA, 37mm., 0.8 micron cellulose ester  
monitor" - obtainable from the Millipore Cor-  
p.

may be obtained from the W. H. Jelly and Co.,  
131.

ined from: a) Mine Safety Appliance Co., Pitts-  
-illson Products, Div., Reading, Penna. 19603;  
ruments Inc., Fall River, Mass. 02720.

#### 4.6 Miscellaneous:

A 1:1 mixture of dimethyl phthalate and diethyl oxalate is needed  
for preparing the sample mounting medium.

Tweezers, scalpel, lens tissue, and a camel's hair brush for maintain-  
ing cleanliness of the microscope and accessories are needed.

#### 5. CLEANLINESS:

- 5.1 All equipment used in fiber counting should be isolated and used  
only for this purpose. Fiber counting should be done in the cleanest  
part of the laboratory; and, if possible, in a separate room. The  
microscope and accessories should be kept clean at all times. Pro-  
cedures for care of the microscope and accessories should follow the  
manufacturer's recommendations.

#### 6. PREPARATION OF MOUNTING SOLUTION:

- 6.1 The mounting solution used in this method is a 1:1 mixture of  
dimethyl phthalate and diethyl oxalate. This solution may be used  
when the sample is to be counted within 24 hours after preparation.  
If samples are to stand for longer periods of time an alternate solution  
should be used. This solution may be prepared by dissolving 0.05  
grams of membrane filter material per milliliter of solution. The pur-  
pose of the dissolved material is to provide a solution with as high a  
viscosity as possible without being difficult to handle. The highly  
viscous solution delays the migration of particles outward from  
the center of the mount. The samples will remain stable with no  
apparent loss in concentration for approximately 30 days.

#### 7. AEROSOL FIELD MONITOR AND FILTERS:

- 7.1 The monitors may be reused after each test by cleaning them thor-  
oughly in warm detergent water and rinsing with distilled water.

#### 8. CALIBRATION OF EQUIPMENT:

##### 8.1 Pump:

The frequency of calibration is dependent on the use, care, and  
handling of the instrument. The calibration for the suction pump  
is carried out with a filter in place. It is suggested that a primary or  
intermediate standard be used in the calibration.

- 8.1.1 Suction pumps equipped with a floating-ball gauge may be  
calibrated by a primary standard. The set up for this pro-  
cedure is shown in Fig. 1. A primary standard provides a  
direct measure of air volume displaced per unit time.

Before starting the calibration, be sure the pump battery is fully charged. Fill a 250 ml. graduate with water and invert in a beaker of water. Note the starting level in the graduate. Let the water come to room temperature. Adjust the ball float to one of the lines on the gauge scale. Insert the tubing from the discharge side of the pump into the graduate. Start the pump and timer simultaneously. As the liquid level in the graduate approaches the liquid level in the beaker, simultaneously stop the timer and lift the graduate from the discharge tube. The pump is still operating. The mouth of the graduate should remain below the level of water in the beaker during this step. Stop the pump and note the final level in the graduate.

Calculate the flow rate in standard liters/min. from the formula,

$$Q = \frac{V_i - V_f}{T} \times \frac{60}{1000} \times K$$

where:

- Q = flow rate in standard liters/min.
- V<sub>i</sub> = initial volume in ml. from graduate.
- V<sub>f</sub> = final volume in ml. from graduate.
- T = time in seconds of calibration.
- 60 = conversion from seconds to minutes when T is in seconds.
- 1000 = conversion from ml. to liters.
- K = factor taking into consideration — temperature and pressure corrections.

To calculate the value of K, use the following formula,

$$K = \frac{P_a - P_v}{760} \times \frac{530}{460 + F}$$

where:

- P<sub>a</sub> = atmospheric pressure in mmHg.
- P<sub>v</sub> = vapor pressure of water at room temperature in mmHg.
- F = room temperature in °F

Duplicate tests should be run at each major scale division on the flowmeter gauge. On a plot of flowmeter setting

versus air flow rate (Q), draw the points. From this line, drawing to two liters/min. is adjusted to this setting.

An alternate method for calibrating the test meter, which is an indirect method, the pump's suction displacement of water in the meter is indicated by the meter. After the meter is leveled and the apparatus is assembled as shown, time the wet test meter. If the meter does not indicate proper setting, adjustment should be made. The actual air flow times a conversion factor (K). The calculation of the value of K is shown above in the primary calibration.

## 8.2 Microscope:

Before using the microscope for counting, the microscope must be in proper adjustment. This procedure is as follows:

- 8.2.1 Screw the 10x and 43x phase contrast nosepiece and insert the 10x objective.
- 8.2.2 Position the 10x objective on the slide on the stage, turn on the diaphragm on the condenser and focus.
- 8.2.3 Set the phase ring to the correct position.
- 8.2.4 Close the iris diaphragm on the condenser to a point where only a small bright spot is visible. Adjust the iris set screws on the condenser so the spot is centered in the field.
- 8.2.5 Focus sides of Decagon with the condenser. Sides will appear blue when focused down. The proper focus is when the diaphragm to cover entire field of view.
- 8.2.6 With the phase ring setting still correct, insert the centering

he calibration, be sure the pump battery is  
 ll a 250 ml. graduate with water and invert  
 ater. Note the starting level in the graduate.  
 oms room temperature. Adjust the ball  
 e lib... on the gauge scale. Insert the tubing  
 rge side of the pump into the graduate.  
 and timer simultaneously. As the liquid  
 duate approaches the liquid level in the  
 ously stop the timer and lift the graduate  
 ge tube. The pump is still operating. The  
 aduate should remain below the level of  
 ker during this step. Stop the pump and  
 l in the graduate.

w rate in standard liters/min. from the

$$\frac{V_i - V_f}{T} \times \frac{60}{1000} \times K$$

low rate in standard liters/min.

initial volume in ml. from graduate.

final volume in ml. from graduate.

time in seconds of calibration.

conversion from seconds to minutes when

is in seconds.

conversion from ml. to liters.

factor taking into consideration — tempera-

ture & pressure corrections.

value of K, use the following formula,

$$\frac{P_a - P_v}{760} \times \frac{530}{460 + F}$$

atmospheric pressure in mmHg.

apor pressure of water at room tempera-

ture in mmHg.

room temperature in °F

ould be run at each major scale division  
 gauge. On a plot of flowmeter setting

versus air flow rate (Q), draw the best straight line through  
 the points. From this line, the flowmeter setting correspond-  
 ing to two liters/min. is determined. The pump should be  
 adjusted to this setting when taking samples.

An alternate method for calibrating pumps is by using a wet  
 test meter, which is an intermediate standard. In this  
 method, the pump's suction is determined by the displace-  
 ment of water in the meter by the gas being measured. This  
 displacement is indicated by a dial on the face of the meter.  
 After the meter is leveled and the sight gauge properly filled,  
 the apparatus is assembled as shown in Fig. 2. Using a stop-  
 watch, time the wet test meter displacement for one minute.  
 If the meter does not indicate two LPM at 6" H<sub>2</sub>O, then an  
 adjustment should be made on the pump to obtain the  
 proper setting. The actual air flow rate is the calibrated rate  
 times a conversion factor (K) for temperature and pressure.  
 The calculation of the value of K is the same as described  
 above in the primary calibration procedure.

## 8.2 Microscope:

Before using the microscope for counting, its phase contrast optics  
 must be in proper adjustment. This procedure is described below.

- 8.2.1 Screw the 10x and 43x phase contrast objectives into the  
 nosepiece and insert the 10x eyepieces.
- 8.2.2 Position the 10x objective for viewing, place a specimen  
 slide on the stage, turn on the illuminator, and set the iris  
 diaphragm on the condenser about half open.
- 8.2.3 Set the phase ring to the 0 position and focus on the speci-  
 men.
- 8.2.4 Close the iris diaphragm on the illuminator down to the  
 point where only a small bright area is visible in the field.  
 Adjust the iris set screws on the illuminator until this spot  
 is centered in the field.
- 8.2.5 Focus sides of Decagon with phase contrast control knob.  
 Sides will appear blue when focused up and red when focus-  
 ed down. The proper focus is mid-point. Open the iris dia-  
 phragm to cover entire field of view.
- 8.2.6 With the phase ring setting still at 0, remove the right eye-  
 piece and insert the centering telescope. Adjust the tele-

scope until the dark circle comes into sharp focus and lock it in position with the set screw provided.

- 8.2.7 Turn the phase ring to the 10 setting. By using the adjustment screws on the side of the condenser, move the rings until they are perfectly concentric.
- 8.2.8 Rotate the 43x objective into viewing position and turn the phase ring to the 43 setting. Align the two rings as described in 8.2.7.
- 8.2.9 Remove the telescope and replace the 10x eyepiece. Use fine adjustment for proper focus. The microscope is now adjusted for phase contrast viewing at 430x magnification.

### 8.3 Reticle:

The microscope must contain a reticle. Calibration of the reticle is done by means of a stage micrometer. The Porton reticle is a rectangle divided into two squares. The counting area is divided into six rectangles which constitutes the left square. A series of circles in which every other circle doubles in diameter is located on the top and bottom of the large rectangle. The diameter, D, of each is given by the formula:  $D = L\sqrt{2^N}$  where L is a unit of length and N represents the circle number. The length of the large reticle is 200 L units. By measuring the actual length of the rectangle using the stage micrometer and dividing this value by 200, the value of L at 430x magnification is obtained. With this value for L, the actual diameter of each of the circles can be determined.

The calibration of the Porton reticle need be done only once, unless the eyepieces or objectives are changed, the magnification is altered in any way, or the microscope had been disassembled for cleaning prior to use.

### 9. SAMPLING:

- 9.1 The suction pump battery must be in a fully charged condition before each test. Recommended battery life for constant pumping rate is furnished by the manufacturer and must not be exceeded for any one test (see paragraph 4.4).
- 9.2 After inserting pad and filter into each of the plastic holders (Aerosol Field Monitors) they are immediately capped and plugged. The junctures of the holders are then sealed with cellulose bands.

All loaded monitors must be carried to the test sites in a closed carrying case. An extra monitor must be taken along to serve as a blank. All monitors must be marked for identification.

- 9.3 At the test site, the top section of the filter" testing. The outlet plus an appropriate length of 1/4 inch into the pump intake. Sampling can be either in taking general area samples it assembly be mounted on a tripod floor level. For a personal sample the assembly is worn by the person being tested close to the person's breathing zone.

The sampling time should be gauged the filter for satisfactory counting. E avoided since no dilution is possible flow rate of two liters per minute. field monitor must be re-sealed.

### 10. COUNTING:

- 10.1 Remove the cover of the field monitor be examined. Cut out a pie shaped section with a scalpel. The section to be about 1 cm.
- 10.2 Place a drop of mounting solution on the drop into a shape corresponding to the drop.
- 10.3 Remove the sample section from the tweezers and deposit it, sample side up and cover with a clean cover glass. Push with the tweezers until it is in intimate contact with the cover glass. Mounts may take 30 minutes if the solution is completely dissolved.
- 10.4 Place a sample slide on the stage and focus because dust is retained only in the top section. The total fiber count should be at least the number of fibers seen, whichever is less.

If a fiber crosses the limits of the counter, two adjacent sides are counted. The side of the counter, but they are always the same. The fibers counted are estimated as to either the top or bottom of the reticle three size groups: all fibers seen, all fibers longer than ten microns, and all fibers longer than ten microns.

rk circle comes into sharp focus and lock the set screw provided.

g to the 10 setting. By using the adjust-  
ne s of the condenser, move the rings  
actly concentric.

jective into viewing position and turn the  
3 setting. Align the two rings as described

cope and replace the 10x eyepiece. Use  
or proper focus. The microscope is now  
contrast viewing at 430x magnification.

tain a reticle. Calibration of the reticle is  
micrometer. The Porton reticle is a rec-  
squares. The counting area is divided into  
tutes the left square. A series of circles in  
oubles in diameter is located on the top  
rectangle. The diameter,  $D$ , of each is given  
 $2^N$  where  $L$  is a unit of length and  $N$   
er. The length of the large reticle is 200  $L$   
actual length of the rectangle using the  
ding this value by 200, the value of  $L$  at  
ined. With this value for  $L$ , the actual dia-  
can be determined.

ton reticle need be done only once, unless  
are changed, the magnification is altered  
scop and been disassembled for cleaning

y must be in a fully charged condition  
ended battery life for constant pumping  
anufacturer and must not be exceeded for  
1.4).

er into each of the plastic holders (Aerosol  
mediately capped and plugged. The junc-  
n sealed with cellulose bands.

se carried to the test sites in a closed carry-  
must be taken along to serve as a blank. All  
for identification.

10

- 9.3 At the test site, the top section of the monitor is removed for "open filter" testing. The outlet plug is removed, the adapter inserted, and an appropriate length of 1/4 inch tubing is connected between it and the pump intake. Sampling can be either general area or personalized. In taking general area samples it is recommended that the filter assembly be mounted on a tripod approximately five feet above floor level. For a personal sample the entire pump and filter assembly is worn by the person being tested with the filter attached as close to the person's breathing zone as is feasible.

The sampling time should be gauged to obtain sufficient material on the filter for satisfactory counting. Excessive dust deposits should be avoided since no dilution is possible. Sampling shall be at an air flow rate of two liters per minute. When the test is completed the field monitor must be re-sealed.

## 10. COUNTING:

- 10.1 Remove the cover of the field monitor to expose the filter that is to be examined. Cut out a pie shaped section of the filter using a clean scalpel. The section to be about 1 cm. on the arc side.
- 10.2 Place a drop of mounting solution on a clean frosted-end slide. Smear the drop into a shape corresponding to the size of the sample section.
- 10.3 Remove the sample section from the plastic field monitor with clean tweezers and deposit it, sample side up, onto the mounting solution and cover with a clean cover glass. Press lightly on the cover glass with the tweezers until it is in intimate contact with the sample section and solution, being careful not to trap any bubbles under the cover glass. Mounts may take 30-60 minutes before the sample section is completely dissolved.
- 10.4 Place a sample slide on the stage and focus on the top surface of filter because dust is retained only in the top 10 to 15 microns of the filter. The total fiber count should be at least 100 fibers, or twenty fields, whichever is less.

If a fiber crosses the limits of the counting field, only those crossing two adjacent sides are counted. The sides chosen are at the discretion of the counter, but they are always the same for any given counter. The fibers counted are estimated as to length, using the circles at either the top or bottom of the reticle. The counts are recorded in three size groups: all fibers seen, all fibers longer than five microns, and all fibers longer than ten microns.

# PRIMARY CALIBRATION





Fig. No. 1


12

# INTERMEDIATE CALIBRATION




HER 0001584

13




meter



1. No. 1  
2

INTERMEDIATE CALIBRATION



13

HER 0001585

Fig. No. 2

10.5 In shifting from one area of examination to another, the choice must be random. Once chosen, the field must be counted regardless of the number of fibers observed. In shifting from one field to another the operator should take his eyes away from the eyepieces and concentrate them on moving the stage to a new location. The choice of the new location should be arbitrary, except that areas near the edge of the sample should be avoided. During the count it is very important to constantly adjust the fine focus back and forth. This brings into focus fibers that might not be seen otherwise.

# 11. CALCULATION OF RESULTS:

The concentration of fibers shall be expressed as the number of fibers per cubic milliliter of air reported to two (2) significant figures only.

$$\text{Concentration} = \frac{(\text{Av. Fiber Count}) (\text{Filter Area})}{(\text{Field Area}) (\text{Sample Volume}) (1000)}$$

where:

Av. Fiber Count: Average Fiber Count Per Size Range.

Filter Area: 855mm<sup>2</sup> for 37mm. dia. membrane filters.

Field Area: Area of counting field in mm<sup>2</sup>.

Sample Volume: Total Time x Flow Rate, Express in liters.

1000: Converts liters to milliliters.

If blanks are used, subtract the average blank count of the same size range from the average fiber count.

## INDUSTRIAL HYGIENE V

1. STATION NUMBER \_\_\_\_\_ PU  
DEPARTMENT \_\_\_\_\_  
PLANT \_\_\_\_\_ TI  
DIVISION \_\_\_\_\_ TI  
DATE SAMPLED \_\_\_\_\_ EI  
DATE ANALYZED \_\_\_\_\_ V

2. STATION LOCATION \_\_\_\_\_  
3. SPECIFIC PRODUCT AND MATERIALS \_\_\_\_\_  
4. OPERATORS AND PROTECTION \_\_\_\_\_  
5. VENTILATING CONTROL \_\_\_\_\_  
6. DRAFTS \_\_\_\_\_ TOTAL  
7. REMARKS \_\_\_\_\_

ANALYSIS: \_\_\_\_\_ F/cc

TLV: \_\_\_\_\_ F/cc

SAMPLED BY \_\_\_\_\_

TOT. \_\_\_\_\_

AV. \_\_\_\_\_

15

HER 0001586

A

**the average blank count of the same size range**

15 HER 0001587

#### BIBLIOGRAPHY ON MEMBRANE FILTER METHOD

1. Edwards, G. H. and J. R. Lynch: The U. S. Public Health Service Method for Membrane Filter Enumeration of Asbestos Dust, Unpublished, 1966.
2. Ayer, E. A. and J. R. Lynch: Measurement of Dust Exposures in the Asbestos Textile Industry. American Industrial Hygiene Association, Journal 27, September - October 1966.
3. Ayer, H. E., G. H. Edwards, J. J. Fanney, Jr., and J. R. Lynch: Relationship of Impinger Counts to Fiber Concentrations by Membrane Filter in Asbestos Textile Plants. Delivered at the annual meeting of the American Industrial Hygiene Association, Houston, Texas, 1965.
4. Bayer, S. G., T. A. Brown, and R. D. Zumwalde: Equipment and Procedures for Mounting Millipore Filters and Counting Asbestos Fibers by Phase Contrast Microscopy. H.E.W. Public Health Service, February, 1969.
5. Transactions of the Thirtieth Annual Meeting of the American Conference of Governmental Industrial Hygienists. Subcommittee on Threshold Limit Values, St. Louis, Mo., 1968.
6. Weidner, R. B.: Personal Respirable Mass Sampling Procedure, H. E. W. Occupational Health Program, 1968.
7. Technical Note No. 1, Measurement of Airborne Asbestos Dust by the Membrane Filter Method, Asbestos Research Council, England, 1969.
8. Technical Data Note No. 13, Standards for Asbestos Dust Concentration for Use with Asbestos Regulations, Her Majesty's Stationary Office, England, 1969.
9. Hygiene Standards for Chrysotile Asbestos Dust, Committee on Hygiene Standards of the British Occupational Hygiene Society. Published in Ann. Occupational Hygiene Volume II, 1968.