

Name: Morgan  
Next test due:

TSI MODEL 8015 PORTACOUNT FIT-TESTING SOFTWARE  
Version 2.01 S/N 927  
VISTA POLYMERS

FIT TEST REPORT

-----  
Test date: 08-03-1994  
Test time: 1:20:57 pm  
Test subject last name: Morgan  
Test subject first name: Erskin  
ID number: 587881971  
Operator name: H. Pruitt  
Next fit test due:  
Test agent: Ambient particles  
Test device: TSI MODEL 8010 PORTACOUNT

RESPIRATOR ID

Size: Medium  
Model: Ultravue Facepiece  
Manufacturer: MSA  
Approval number: TC-19C-113

Additional notes:  
Additional notes:

TEST DATA

-----  
Pass/Fail level: 100  
Test cycles.....: 1

| Ex. | Ambient<br>(Part/cc) | Mask<br>(Part/cc) | Fit Factor | Pass/Fail |
|-----|----------------------|-------------------|------------|-----------|
| NB  | 16000                | 0.210             | 76400      | PASS      |
| DB  | 16400                | 0.720             | 22800      | PASS      |
| SS  | 16950                | 0.260             | 65200      | PASS      |
| UD  | 16550                | 0.140             | 118000     | PASS      |
| R   | 16450                | 15.600            | 1050       | PASS      |
| NB  | 16350                | 0.060             | 273000     | PASS      |

OVERALL FIT FACTOR = 5787 PASS  
(Calculated per ANSI Z88.2-1980)

Operator H. Pruitt Date 8-3-94  
Test Subject Erskin Morgan Date 8-3-94