

Material: B-D, 20 ml. Vacutainers
Source: See letter of October 23 from J. J. Kleiner,
Becton-Dickinson Company
Required: Lead contamination study

Ten tubes of a lot of 114 were handled as in our previous test
with the following results.

Microgram of Pb Found

0.8
0.6
0.7
0.4
0.1
0.6
0.2
0.2
0.2
0.3

Avg. 0.42

Reported: 10-26-45

N13970

DE 222222