

Name: JACKSON
Next test due: .

TSI MODEL 8015 PORTACOUNT FIT-TESTING SOFTWARE
Version 2.01 S/N 927
VISTA POLYMERS

FIT TEST REPORT

Test date: 10-01-1994
Test time: 6:39:27 pm
Test subject last name: JACKSON
Test subject first name: ROBERT
ID number: 428924010
Operator name: STEWART
Next fit test due:
Test agent: Ambient particles
Test device: TSI MODEL 8010 PORTACOUNT

RESPIRATOR ID

Size: LARGE
Model: ULTRAVUE FACEPIECE
Manufacturer: 3M
Approval number: TC-21C-488

Additional notes:
Additional notes:

TEST DATA

Pass/Fail level: 100
Test cycles.....: 1

Ex.	Ambient (Part/cc)	Mask (Part/cc)	Fit Factor	Pass/Fail
NB	6585	1.830	3600	PASS
DB	6760	0.900	7510	PASS
SS	7245	6.630	1090	PASS
UD	7485	1.500	4990	PASS
R	7245	7.470	970	PASS
NB	6895	19.400	353	PASS

OVERALL FIT FACTOR = 1113 PASS
(Calculated per ANSI Z88.2-1980)

Operator Arthur J. Stewart Date 10-1-94

Test Subject Robert A. Jackson Date 10-1-94