

Name: Morgan
Next test due:

TSI MODEL 8015 PORTACOUNT FIT-TESTING SOFTWARE
Version 2.01 S/N 927
VISTA POLYMERS

FIT TEST REPORT

Test date: 08-08-1994
Test time: 8:31:14 am
Test subject last name: Morgan
Test subject first name: Jimmie
ID number: 428024556
Operator name: H Pruitt
Next fit test due:
Test agent: Ambient particles
Test device: TSI MODEL 8010 PORTACOUNT

RESPIRATOR ID

Size: Medium
Model: Ultravue
Manufacturer: Facepiece
Approval number:

Additional notes:
Additional notes:

TEST DATA

Pass/Fail level: 100
Test cycles.....: 1

Ex.	Ambient (Part/cc)	Mask (Part/cc)	Fit Factor	Pass/Fail
NB	15650	0.100	157000	PASS
DB	15500	0.040	388000	PASS
SS	16000	0.200	80100	PASS
UD	16050	0.230	70000	PASS
R	15650	0.160	98100	PASS
NB	16400	0.140	117000	PASS

OVERALL FIT FACTOR = 110178 PASS
(Calculated per ANSI Z88.2-1980)

Operator Helen Pruitt Date 8-8-94
Test Subject Jimmie A. Morgan Date 8-8-94