

Name: Morgan
Next test due:

TSI MODEL 8015 PORTACOUNT FIT-TESTING SOFTWARE
Version 2.01 S/N 927
VISTA POLYMERS

FIT TEST REPORT

Test date: 08-12-1994
Test time: 7:37:02 am
Test subject last name: Morgan
Test subject first name: Robert
ID number: 225700771
Operator name: S. Carter
Next fit test due:
Test agent: Ambient particles
Test device: TSI MODEL 8010 PORTACOUNT

RESPIRATOR ID

Size: Medium
Model: Ultravue Facepiece
Manufacturer: MSA
Approval number: TC-19C-113

Additional notes:
Additional notes:

TEST DATA

Pass/Fail level: 100
Test cycles.....: 1

Ex.	Ambient (Part/cc)	Mask (Part/cc)	Fit Factor	Pass/Fail
NB	30650	0.500	61300	PASS
DB	30850	0.150	205000	PASS
SS	30750	1.420	21600	PASS
UD	30850	1.530	20100	PASS
R	30550	6.940	4400	PASS
NB	30000	0.100	300000	PASS

OVERALL FIT FACTOR = 17249 PASS
(Calculated per ANSI Z88.2-1980)

Operator Shelby D. Carter Date 8-12-94
Test Subject Robert Morgan Date 8-12-94