

Name: Noland
Next test due:

TSI MODEL 8015 PORTACOUNT FIT-TESTING SOFTWARE
Version 2.01 S/N 927
VISTA POLYMERS

FIT TEST REPORT

Test date: 10-12-1994
Test time: 8:12:14 pm
Test subject last name: Noland
Test subject first name: Terry
ID number: 416-70-7496
Operator name: Pruitt
Next fit test due:
Test agent: Ambient particles
Test device: TSI MODEL 8010 PORTACOUNT

RESPIRATOR ID

Size: SMALL
Model: Ultravue Facepiece
Manufacturer: 3M
Approval number: TC-21C-488

Additional notes:
Additional notes:

TEST DATA

Pass/Fail level: 100
Test cycles.....: 1

Ex.	Ambient (Part/cc)	Mask (Part/cc)	Fit Factor	Pass/Fail
NB	2005	0.290	6930	PASS
DB	1990	0.490	4070	PASS
SS	1935	0.160	12100	PASS
UD	1880	0.170	11000	PASS
R	1755	4.110	427	PASS
NB	1690	0.770	2200	PASS

OVERALL FIT FACTOR = 1786 PASS
(Calculated per ANSI Z88.2-1980)

Operator Helen Pruitt Date 10-12-94

Test Subject Terry L. Noland Date 10-12-94