

Name: McWhirter  
Next test due:

TSI MODEL 8015 PORTACOUNT FIT-TESTING SOFTWARE  
Version 2.01 S/N 927  
VISTA POLYMERS

## FIT TEST REPORT

-----  
Test date: 08-01-1994  
Test time: 12:26:34 pm  
Test subject last name: McWhirter  
Test subject first name: John W.  
ID number: 427944939  
Operator name: S. Carter  
Next fit test due:  
Test agent: Ambient particles  
Test device: TSI MODEL 8010 PORTACOUNT

## RESPIRATOR ID

Size: Medium  
Model: Ultravue Facepiece  
Manufacturer: MSA  
Approval number: TC-19C-113

Additional notes:  
Additional notes:

## TEST DATA

-----  
Pass/Fail level: 100  
Test cycles.....: 1

| Ex. | Ambient<br>(Part/cc) | Mask<br>(Part/cc) | Fit Factor | Pass/Fail |
|-----|----------------------|-------------------|------------|-----------|
| NB  | 26250                | 0.450             | 58500      | PASS      |
| DB  | 24300                | 6.880             | 3540       | PASS      |
| SS  | 22150                | 4.110             | 5400       | PASS      |
| UD  | 21800                | 3.070             | 7120       | PASS      |
| R   | 21100                | 20.000            | 1050       | PASS      |
| NB  | 21100                | 1.170             | 18000      | PASS      |

OVERALL FIT FACTOR = 3674 PASS  
(Calculated per ANSI Z88.2-1980)

Operator Shelby Carter Date 8-1-94

Test Subject John McWhirter Date 8-1-94