

VCM CONTINUOUS MONITORING LOG SHEET GC NO. 4 - LABORATORY

4-6-92

NAME:	TALLIE																								WJ					
TIME	0700	0800	0900	1000	1100	1200	1300	1400	1500	1600	1700	1800	1900	2000	2100	2200	2300	2400	0100	0200	0300	0400	0500	0600						
Stream No.	1	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0						
	2	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0						
	3	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0						
	4	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0						
	5	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0						
	6	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0						
	7	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0						
	8	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0						
	9	0	0	7.2	0	0	0	0	0	0	0	0	0	10.0	10.0	0	0	0	0	0	0	0	0	0						
	10	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0						
Flame On		/	/	/	/	/	/	/	/	/	/	/	/	/	/	/	/	/	/	/	/	/	/	/						
Plant Air On.		/	/	/	/	/	/	/	/	/	/	/	/	/	/	/	/	/	/	/	/	/	/	/						

GC No. 4 - LAB

Stream Locations:

- 1 Front Guard House
- 2 North End VCM Railcar Unloading Station
- 3 South End VCM Railcar Unloading Station
- 4 Quality Control Lab
- 5 GC Lab
- 6 South of P1 Building
- 7 Compound Line 1
- 8 D-700 Dump System
- 9 D-745 Dump System
- 10 Plant Breathing Air

0830 to 3 9.0

2000-2200 - 10.0 - 745 dump system
called Bill King, + informed.