

Name: Sykes
Next test due:

TSI MODEL 8015 PORTACOUNT FIT-TESTING SOFTWARE
Version 2.01 S/N 927
VISTA POLYMERS

FIT TEST REPORT

Test date: 08-12-1994
Test time: 12:19:10 pm
Test subject last name: Sykes
Test subject first name: Jeffrey
ID number: 587278743
Operator name: S. Carter
Next fit test due:
Test agent: Ambient particles
Test device: TSI MODEL 8010 PORTACOUNT

RESPIRATOR ID

Size: Large
Model: Ultravue Facepiece
Manufacturer: MSA
Approval number: TC-19C-113

Additional notes:
Additional notes:

TEST DATA

Pass/Fail level: 100
Test cycles.....: 1

Ex.	Ambient (Part/cc)	Mask (Part/cc)	Fit Factor	Pass/Fail
NB	31450	0.210	150000	PASS
DB	31400	0.290	108000	PASS
SS	31650	0.270	117000	PASS
UD	31500	0.180	175000	PASS
R	31250	3.690	8480	PASS
NB	31650	0.190	166000	PASS

OVERALL FIT FACTOR = 38927 PASS
(Calculated per ANSI Z88.2-1980)

Operator Shelby D. Carter Date 8-12-94
Test Subject Jeffrey P. Sykes Date 8-12-94