

Name: Wolf
Next test due:

TSI MODEL 8015 PORTACOUNT FIT-TESTING SOFTWARE
Version 2.01 S/N 927
VISTA POLYMERS

FIT TEST REPORT

Test date: 08-08-1994
Test time: 11:23:10 am
Test subject last name: Wolf
Test subject first name: Gary
ID number: 423116466
Operator name: H Pruitt
Next fit test due:
Test agent: Ambient particles
Test device: TSI MODEL 8010 PORTACOUNT

RESPIRATOR ID

Size: Small
Model: Ultravue
Manufacturer: Facepiece
Approval number:

Additional notes:
Additional notes:

TEST DATA

Pass/Fail level: 100
Test cycles.....: 1

Ex.	Ambient (Part/cc)	Mask (Part/cc)	Fit Factor	Pass/Fail
NB	14900	0.590	25300	PASS
DB	15450	0.110	140000	PASS
SS	15850	0.290	54800	PASS
UD	15850	0.090	176000	PASS
R	16000	1.400	11400	PASS
NB	15950	0.280	57000	PASS

OVERALL FIT FACTOR = 34118 PASS
(Calculated per ANSI Z88.2-1980)

Operator

Helen Pruitt

Date

8-8-94

Test Subject

[Signature]

Date

8/8/94