

Name: Sanders
Next test due:

TSI MODEL 8015 PORTACOUNT FIT-TESTING SOFTWARE
Version 2.01 S/N 927
VISTA POLYMERS

FIT TEST REPORT

Test date: 08-08-1994
Test time: 9:56:11 am
Test subject last name: Sanders
Test subject first name: Gary
ID number: 587158195
Operator name: H Pruitt
Next fit test due:
Test agent: Ambient particles
Test device: TSI MODEL 8010 PORTACOUNT

RESPIRATOR ID

Size: Medium
Model: Ultravue
Manufacturer: Facepiece
Approval number:

Additional notes:
Additional notes:

TEST DATA

Pass/Fail level: 100
Test cycles.....: 1

Ex.	Ambient (Part/cc)	Mask (Part/cc)	Fit Factor	Pass/Fail
NB	12900	1.750	7390	PASS
DB	12950	1.220	10600	PASS
SS	12900	0.590	21900	PASS
UD	12750	0.430	29700	PASS
R	12750	4.500	2840	PASS
NB	12600	0.270	46800	PASS

OVERALL FIT FACTOR = 8792 PASS
(Calculated per ANSI Z88.2-1980)

Operator Helen Pruitt Date 8-8-94
Test Subject Gary W. Sanders Date 8-8-94