

Name: White
Next test due:

TSI MODEL 8015 PORTACOUNT FIT-TESTING SOFTWARE
Version 2.01 S/N 927
VISTA POLYMERS

FIT TEST REPORT

Test date: 08-12-1994
Test time: 10:47:31 am
Test subject last name: White
Test subject first name: Willie W.
ID number: 587016028
Operator name: S. Carter
Next fit test due:
Test agent: Ambient particles
Test device: TSI MODEL 8010 PORTACOUNT

RESPIRATOR ID
Size: Medium
Model: Ultravue Facepiece
Manufacturer: MSA
Approval number: TC-19C-113

Additional notes:
Additional notes:

TEST DATA

Pass/Fail level: 100
Test cycles.....: 1

Ex.	Ambient (Part/cc)	Mask (Part/cc)	Fit Factor	Pass/Fail
NB	29200	0.020	1460000	PASS
DB	28950	2.800	10300	PASS
SS	28950	0.830	34900	PASS
UD	28950	0.160	181000	PASS
R	29850	3.440	8700	PASS
NB	29700	0.220	135000	PASS

OVERALL FIT FACTOR = 23594 PASS
(Calculated per ANSI Z88.2-1980)

Operator Shelby D. Carter Date 8-12-94
Test Subject Willie W. White Date 8-12-94