

Name: HAYNES
Next test due:

TSI MODEL 8015 PORTACOUNT FIT-TESTING SOFTWARE
Version 2.01 S/N 927
VISTA POLYMERS

FIT TEST REPORT

Test date: 09-30-1994
Test time: 9:08:31 pm
Test subject last name: HAYNES
Test subject first name: COLEMAN
ID number: 587184282
Operator name: STEWART
Next fit test due:
Test agent: Ambient particles
Test device: TSI MODEL 8010 PORTACOUNT

RESPIRATOR ID
Size: MEDIUM
Model: ULTRAVUE FACEPIECE
Manufacturer: 3M
Approval number: TC-31C-488

Additional notes:
Additional notes:

TEST DATA

Pass/Fail level: 100
Test cycles.....: 1

Ex.	Ambient (Part/cc)	Mask (Part/cc)	Fit Factor	Pass/Fail
NB	13200	0.820	16100	PASS
DB	14050	0.930	15100	PASS
SS	14550	2.240	6510	PASS
UD	14300	2.010	7140	PASS
R	14150	15.900	889	PASS
NB	13650	1.950	7000	PASS

OVERALL FIT FACTOR = 3551 PASS
(Calculated per ANSI Z88.2-1980)

Operator Arthur J. Stewart Date 9-30-94

Test Subject Coleman Haynes Date 9-30-94