

Name: Uptain
Next test due:

TSI MODEL 8015 PORTACOUNT FIT-TESTING SOFTWARE
Version 2.01 S/N 927
VISTA POLYMERS

FIT TEST REPORT

Test date: 08-12-1994
Test time: 1:21:37 pm
Test subject last name: Uptain
Test subject first name: Jerald
ID number: 418542511
Operator name: S. Carter
Next fit test due:
Test agent: Ambient particles
Test device: TSI MODEL 8010 PORTACOUNT

RESPIRATOR ID

Size: Medium
Model: Ultravue Facepiece
Manufacturer: MSA
Approval number: TC-19C-113

Additional notes:
Additional notes:

TEST DATA

Pass/Fail level: 100
Test cycles.....: 1

Ex.	Ambient (Part/cc)	Mask (Part/cc)	Fit Factor	Pass/Fail
NB	37450	3.690	10100	PASS
DB	37000	3.370	10900	PASS
SS	35950	1.320	27200	PASS
UD	34800	2.260	15400	PASS
R	33700	6.120	5520	PASS
NB	32400	1.410	23000	PASS

OVERALL FIT FACTOR = 11603 PASS
(Calculated per ANSI Z88.2-1980)

Operator Shelby D. Carter Date 8-12-94
Test Subject J. J. C. K. Date 8-12-94