PLAINTIFF'S EXHIBIT

ATT-134

Bell Laboratories

600 Mountain Avenue
Murray Hill, New Jersey 07974
Phone (201) 582-3000

August 4, 1978

E. R. Sullivan
ASTM
1916 Race Street
Philadelphia, PA 19103

Dear Mr. Sullivan:

In reviewing the method of testing for asbestos fibers in work place atmosphere, there seems to be a conflicting statement in Item 11.2.2.1. We appear to be recommending both a 3 to 1 length/width ratio and a ratio between 10 to 1 and 20 to 1. This point should be clarified in the final practices. Section 12: the units on Items C, F and B are not clear. It is assumed that you mean micrometers rather than meters.

Sincerely,

W. J. Schreibeis

LLA 000898