

Name: Hamilton
Next test due:

TSI MODEL 8015 PORTACOUNT FIT-TESTING SOFTWARE
Version 2.01 S/N 927
VISTA POLYMERS

FIT TEST REPORT

Test date: 08-21-1994
Test time: 9:17:50 pm
Test subject last name: Hamilton
Test subject first name: Robert J.
ID number: 427869680
Operator name: S. Carter
Next fit test due:
Test agent: Ambient particles
Test device: TSI MODEL 8010 PORTACOUNT

RESPIRATOR ID

Size: M
Model: Ultravue Facepiece
Manufacturer: MSA
Approval number: TC-19C-113

Additional notes:
Additional notes:

TEST DATA

Pass/Fail level: 100
Test cycles.....: 1

Ex.	Ambient (Part/cc)	Mask (Part/cc)	Fit Factor	Pass/Fail
NB	34950	2.230	15600	PASS
DB	34950	1.740	20100	PASS
SS	34300	1.000	34300	PASS
UD	34000	1.120	30400	PASS
R	34150	3.160	10800	PASS
NB	33300	0.710	46900	PASS

OVERALL FIT FACTOR = 20703 PASS
(Calculated per ANSI Z88.2-1980)

Operator Shelby S. Carter Date 8-21-94

Test Subject Robert J. Hamilton Date 8-21-94