

25-1221

|               |                                |              |
|---------------|--------------------------------|--------------|
| J. C. McCool  | D/0020, Bldg. 5-H              |              |
| E. R. Gleason | D/8506, Brecksville R&D Center | Dec. 6, 1974 |

The MSA efficiency samples received November 18 (log 42) have been analyzed for acrylonitrile. The results are summarized in the table below.

| Sample No. | Mg VCN on tube | Mg VCN recovered | % Recovery |
|------------|----------------|------------------|------------|
| MC-35      | 1.35           | 0.85             | 63         |
| 36         | 1.85           | 1.3              | 70         |
| 37         | 1.78           | 1.3              | 73         |
| 38         | 1.82           | 1.3              | 71         |
| 39         | 1.36           | 0.95             | 70         |
| 40         | 1.55           | 1.2              | 77         |
| 41         | 1.83           | 1.5              | 82         |
| 42         | 1.21           | 0.96             | 79         |
| 43         | 1.14           | 0.82             | 72         |

*Elizabeth R. Gleason*  
ELIZABETH R. GLEASON

cad  
cc: R&D Files

8502

23794001

BFG27255