



walter c. mcCrone associates, inc.

CONSULTING: ULTRAMICROANALYSIS • MICROSCOPY • SMALL PARTICLE PROBLEMS • SOLID-STATE CHEMISTRY

27 April 1984

Dr. J. P. Simko, Jr.
Sr. Section Head, Analytical
Department
Colgate-Palmolive Company
909 River Road
Piscataway, New Jersey 08854

Dear Dr. Simko:

This report concerns the asbestos analysis by TEM of six talc samples submitted under your P.O. 634617. Chrysotile asbestos was detected in samples 0613BK, 2713EX and 1813AX. No asbestos was detected in samples S9073, A122831 and A113031. Since this technique is not quantitative, no asbestos concentration can be assigned.

Thank you for consulting McCrone Associates. If you have any questions, please do not hesitate to call.

Sincerely,

Deborah P. Palenik
Electron Microscopist

DPP:fe
Ref: 11871