

0221001
TP003015

Garfield Water Department

Client Sample ID: GL ELMWOOD PARK PUMP STATION

HPLC

Lot-Sample #....: D8C060309-001 Work Order #....: KH51C1AA Matrix.....: WATER
Date Sampled....: 03/05/08 09:00 Date Received...: 03/06/08
Prep Date.....: 03/10/08 Analysis Date...: 03/12/08
Prep Batch #....: 8070135 Analysis Time...: 22:21
Dilution Factor: 1
Method.....: DEN -LC-0012

PARAMETER	RESULT	REPORTING LIMIT	UNITS
PFOS	0.021	0.010	ug/L
PFOA	0.038	0.015	ug/L

SURROGATE	PERCENT RECOVERY	RECOVERY LIMITS
13C4 PFOA	67	(50 - 200)
13C4 PFOS	41 *	(50 - 200)

NOTE(S):

* Surrogate recovery is outside stated control limits.

EXECUTIVE SUMMARY - Detection Highlights

D8C060309

<u>PARAMETER</u>	<u>RESULT</u>	<u>REPORTING LIMIT</u>	<u>UNITS</u>	<u>ANALYTICAL METHOD</u>
GL ELMWOOD PARK PUMP STATION 03/05/08 09:00 001				
PFOS	0.021	0.010	ug/L	DEN -LC-0012
PFOA	0.038	0.015	ug/L	DEN -LC-0012