

Name: Nolen
Next test due:

TSI MODEL 8015 PORTACOUNT FIT-TESTING SOFTWARE
Version 2.01 S/N 927
VISTA POLYMERS

FIT TEST REPORT

Test date: 08-01-1994
Test time: 1:48:16 pm
Test subject last name: Nolen
Test subject first name: David
ID number: 416705064
Operator name: S. Carter
Next fit test due:
Test agent: Ambient particles
Test device: TSI MODEL 8010 PORTACOUNT

RESPIRATOR ID
Size: Medium
Model: Ultravue Facepiece
Manufacturer: MSA
Approval number: TC-19C-113

Additional notes:
Additional notes:

TEST DATA

Pass/Fail level: 100
Test cycles.....: 1

Ex.	Ambient (Part/cc)	Mask (Part/cc)	Fit Factor	Pass/Fail
NB	2145	2.990	718	PASS
DB	1840	0.630	2920	PASS
SS	1695	0.160	10600	PASS
UD	1570	0.100	15700	PASS
R	1430	1.180	1210	PASS
NB	1300	0.240	5430	PASS

OVERALL FIT FACTOR = 2066 PASS
(Calculated per ANSI Z88.2-1980)

Operator Shelby Carter Date 8-1-94
Test Subject David Nolen Date _____