

F. M. Zakrski

Brecksville

J. D. Fannin

Louisville

11/10/75

NOT RECORDED

We are currently fabricating a connection for flasher tubes to sample gases within a person's air line mask. Concern is present over humidity in exhaled air and the thinking is to condition the sample flow through a desiccant, probably silica gel, prior to exposure to the flasher tube.

Can you tell us if silica gel will interfere with the determination?

J. D. Fannin

sjm

cc: R. W. Toole

R. L. Martin