

Name: Thompson
Next test due:

TSI MODEL 8015 PORTACOUNT FIT-TESTING SOFTWARE
Version 2.01 S/N 927
VISTA POLYMERS

FIT TEST REPORT

Test date: 08-03-1994
Test time: 11:38:57 am
Test subject last name: Thompson
Test subject first name: Billy
ID number: 587182925
Operator name: H Pruitt
Next fit test due:
Test agent: Ambient particles
Test device: TSI MODEL 8010 PORTACOUNT

RESPIRATOR ID

Size: Medium
Model: Ultravue Facepiece
Manufacturer: MSA
Approval number: TC-19C-113

Additional notes:
Additional notes:

TEST DATA

Pass/Fail level: 100
Test cycles.....: 1

Ex.	Ambient (Part/cc)	Mask (Part/cc)	Fit Factor	Pass/Fail
NB	18050	0.310	58300	PASS
DB	17850	1.460	12200	PASS
SS	17700	0.290	61200	PASS
UD	17450	1.190	14600	PASS
R	17200	1.430	12000	PASS
NB	17150	0.310	55500	PASS

OVERALL FIT FACTOR = 21030 PASS
(Calculated per ANSI Z88.2-1980)

Operator Helen Pruitt Date 8-3-94

Test Subject Billy Thompson Date 8-3-94