

Name: Strickland
Next test due:

TSI MODEL 8015 PORTACOUNT FIT-TESTING SOFTWARE
Version 2.01 S/N 927
VISTA POLYMERS

FIT TEST REPORT

Test date: 08-12-1994
Test time: 9:50:52 am
Test subject last name: Strickland
Test subject first name: Rodney
ID number: 587706891
Operator name: S. Carter
Next fit test due:
Test agent: Ambient particles
Test device: TSI MODEL 8010 PORTACOUNT

RESPIRATOR ID

Size: Large
Model: Ultravue Facepiece
Manufacturer: MSA
Approval number: TC-19C-113

Additional notes:
Additional notes:

TEST DATA

Pass/Fail level: 100
Test cycles.....: 1

Ex.	Ambient (Part/cc)	Mask (Part/cc)	Fit Factor	Pass/Fail
NB	37500	0.150	250000	PASS
DB	37700	0.640	58900	PASS
SS	38650	0.070	553000	PASS
UD	38150	0.210	181000	PASS
R	37600	2.650	14200	PASS
NB	37350	0.140	266000	PASS

OVERALL FIT FACTOR = 58541 PASS
(Calculated per ANSI Z88.2-1980)

Operator Shelby D. Carter Date 8-12-94

Test Subject Rodney B. Strickland¹ Date 8-12-94