

HYDROCARBON TECHNOLOGY INC.

CUSTOMER: VISTA
WESTLAKE, LA

DATE: 10-23-89
ORDER: 892396
P.O.#: 0383X12224

PRIMARY GAS STANDARD FOR INSTRUMENT CALIBRATION (+OR- 1 %)

CYLINDER NUMBER: CC074029
CGA VALVE TYPE: 350

COMPONENT:

CONCENTRATION:
(MOLAR)

* VINYL CHLORIDE	0.502 PPM	+OR- 0.005 PPM ABSOLUTE
ETHYLENE DICHLORIDE	0.440 PPM	+OR- 0.004 PPM ABSOLUTE
NITROGEN	BALANCE	

* VINYL CHLORIDE VALUE NBS TRACEABLE VIA SRM

CYLINDER PRESSURE: 2000 PSIA
CYLINDER CONTENTS: 135 SCF

ONE YEAR SHELF LIFE FROM DATE OF MANUFACTURE:

MAKE DATE: 10-23-89
EXP. DATE: 10-23-90



ANALYST

VVV 000017858

2621 INDUSTRIAL DRIVE
P.O. BOX 1338 • SULPHUR, LOUISIANA 70664
(318) 882-1701