

"TEST START DATE: 06-19-1989"
 "TEST START TIME: ----"
 " TEST LOCATION: T&D WAREHOUSE"
 " EMPLOYEE NAME: STAN DRAYTON"
 "EMPLOYEE NUMBER: 185-46-9997"
 " DEPARTMENT: T&D"
 "COMMENT FIELD 1: TOW MOTOR OPERATOR"
 "COMMENT FIELD 2: NORMAL DAY"
 " NUMERIC CODES: ---- - - - - - - - - - -"

METROSONICS db-308 SN 2528 V2.3
 REPORT PRINTED 06/19/89 @ 17:06:22

DOUBLING RATE: 3dB FILTER: A WGT
 DOSE CRITERION: 90dB RESPONSE: SLOW
 PRE-CALIBRATION TIME: 6/18/89 @ 19:28:44
 PRE-CALIBRATION RANGE: 43.2dB TO 139.2dB

CALIBRATOR TYPE & SERIAL # : -----

CALIBRATOR CALIBRATION DATE: -----

TEST BEGAN 6/18/89 @ 20:05:51
 TEST LENGTH: 0DAYS 07:20:41

Lav = 87.8dB
 SEL = 131.9dB
 Lmax = 111.3dB ON 6/18/89 @ 20:48:28
 Lpk = 133dB ON 6/19/89 @ 0:36:32
 TIME OVER 115dB 0D 0:00:00.00

8 HR DOSE (80dB CUTOFF) = 54.57%
 8 HR PROJ. DOSE (80dB CUTOFF) = 59.43%
 8 HR DOSE (90dB CUTOFF) = 40.02%
 8 HR PROJ. DOSE (90dB CUTOFF) = 43.59%

Sound Level Meter Calibration

SLM Mfg. METROSONICS db-308 SN 2528

Pre-Cal					
Calibrator SN	General Radio Company 12185				
Location	T&D Warehouse				
Results					
	125	250	500	1000	2000
dBA				114.2	
dBC					
Initials	Date/Time 6/18/89 7:30 AM				

Post-Cal					
Calibrator SN	General Radio Company 12185				
Location	T&D Warehouse				
Results					
	125	250	500	1000	2000
dBA				114.0	
dBC					
Initials	Date/Time 6/18/89 3:55 PM				