

master 45828
BFG TECHNICAL DOCUMENT

BFGoodrich
Chemical Division

RESEARCH AND DEVELOPMENT REPORT

Vinyl Chloride Monomer Concentration in
Working Environments as Related to the Residual
Vinyl Chloride Monomer Concentration in Stripped Latex Products

By
D.E. Weaver - R.E. Shaffer
F. Dwyer

AND FILE NO.

R. J. Meyer /

R. J. Meyer

R. J. Meyer

C. E. Meyer

211A2001

EXECUTIVE SUMMARY

OBJECTIVE

Develop technical information to support a business decision to remove cancer hazard labels from vinyl latex products.

SIGNIFICANT RESULTS

Test data were used to relate ambient vinyl chloride monomer (AVCM) concentrations to residual vinyl chloride monomer (RVCM) levels in extruded products. Worst case conditions were simulated. It was determined that exposure of workers to AVCM concentrations above the OSHA action level is unlikely in real workplace situations.

CONCLUSIONS

In a typical Gess SPM compounding operation the AVCM concentration is estimated to be well below the OSHA action level.

In a typical extrusion operation at Rhineland Paper Company, the AVCM concentration is estimated to be well below the OSHA action level.

In a typical extrusion operation at Sanitas, who use a different extrusion process, the AVCM concentration is estimated to be well below the OSHA action level.

In a typical extrusion operation at the different latex at the different latex, the AVCM concentration is estimated to be well below the OSHA action level.

In a typical extrusion operation at the different latex, the AVCM concentration is estimated to be well below the OSHA action level.

TABLE OF CONTENTS

	Page
EXECUTIVE SUMMARY.....	ii
Objective.....	ii
Significant Results.....	ii
Conclusions.....	ii
Action Taken.....	ii
List of Tables.....	iv
List of Figures.....	v
Summary.....	1
Discussion.....	3
I. Introduction.....	3
II. National Vinyl Chloride Monomer (NVCM) -- Stripped Latex Products.....	4
III. Experimental Plan.....	7
IV. Personnel -- R.E. Shaffer.....	7
V. Sampling -- Sample Tubes -- Test Reports.....	11
VI. Data Analysis.....	24
VII. Summary of Critical Test Data to Date.....	25
VIII. Summary of Critical Test Data to Date.....	25
IX. Summary of Critical Test Data to Date.....	25
X. Summary of Critical Test Data to Date.....	25
XI. Summary of Critical Test Data to Date.....	25
APPENDIX.....	39
List of.....	40
Summary of.....	41
Summary of.....	49
Summary of.....	59
Summary of.....	72
Summary of.....	83
Summary of.....	83

LIST OF TABLES

	<u>Page</u>
I. Plant Data-RVCM, mg/kg - Major Products.....	6
II. Test Data: RVCM and AVCM, Closed Room Experiments.....	12
III. Gas Chromatography Conditions for the Thermal Desorption VCM Analysis.....	16
IV. Chromatography Conditions for VCM Analysis by the CS ₂ Desorption Technique.....	19
V. Blanks: Background VCM Concentration in Air, ppm.....	21
VI. Critical Levels of RVCM in Vinyl Latexes-Closed Room with No Ventilation.....	24
VII. Correlation Constants: Vapor Liquid Equilibria.....	25
VIII. Calculated Critical RVCM Levels in Vinyl Latexes Based on Equilibrium Partial Pressure.....	27
IX. RVCM Calculated from RVCM Loss Data.....	29

LIST OF FIGURES

	<u>Page</u>
1. Schematic Diagram of Experimental Apparatus.....	10
2. AVCM Concentration vs. Latex RVCM Concentration--Raw Data.....	22
3. Measured AVCM Concentration as Related to AVCM Calculated from Latex RVCM Loss During the Test.....	30
4. AVCM vs. Equilibrium VCM Partial Pressure.....	32
5. Ventilation Requirement for OSHA Compliance During a Typical 80% 5% Compounding Operation.....	36

SUMMARY

This report documents technical information that supports a business decision to remove cancer hazard labels from latex products containing vinyl chloride monomer. Controlled experiments in a closed, unventilated room relate ambient concentrations of vinyl chloride monomer (AVCM) to residual vinyl chloride monomer (RVCM) in stripped vinyl latexes. Commercial latex consuming operations have also been analyzed to estimate work place AVCM. Measurements were made by personnel monitoring in one of these operations. Our current standard of 25mg/kg maximum RVCM (wet basis) for vinyl latex products is well within the necessary limit for compliance to standards described by the Occupational Safety and Health Administration (OSHA).

Critical latex RVCM concentrations are defined. At the critical RVCM, the air in a closed, unventilated room will reach the OSHA action level, time weighted average (TWA), 95% confidence, in a 4 hour test period. Test results for three vinyl latexes are:

Latex	Latex Temperature, °C	Critical RVCM, mg/kg
Gom 576	110	22
Gom 450K20	110	30
Gom 352	110	>50

These results are translated to other vinyl latex products based on equilibrium partial pressure data. The assumption is made that diffusion of VC through the latex-air interface is the controlling factor. This is a worst case assumption.

Latex	Electrode, mm	Base Resin	Latex Temperature, °C	Calculated Critical RVCM, mg/kg
Gom 576	1.5	---	110	29
Gom 450K20	1.5	Gom 352	110	29
Gom 352	1.5	Gom 352	110	17

Based on the calculated data, the critical RVCM is approximately 29 mg/kg for the two latexes based on Gom 351 and 17 mg/kg for Gom 352.

Gom 576	30	Gom 352	110	29
Gom 450K20	30	Gom 352	110	29

Obviously, current stripping practices in the industry are sufficient to keep our vinyl latex products below our current standard. The product specifications for Gom 576 and Gom 450K20 could be changed to 15mg/kg and 15mg/kg maximum RVCM, respectively.

A typical Geon 576 compounding operation was modelled. At up to 40mg/kg RVCM, the OSHA action level is not exceeded where ordinary industrial ventilation is provided, viz., 6 turnovers per hour.

A coating operation at Rhineland Paper Co. was analyzed. A clear coating of Geon 576 at 70mg/kg RVCM will not create non compliance AVCM concentrations even if all the RVCM transfers to the air between the coater and the ovens. This again assumes ordinary industrial ventilation. At 22mg/kg RVCM, the AVCM is about 0.10 ppm using the same assumptions.

Both the EFGoodrich Corporate Environmental Health Laboratory and Waples Insurance Companies, a private environmental health engineering service laboratory, confirmed RVCM concentrations at Rhineland to be well under the OSHA action level.

Smith's coats series with two different pigmented Geon 580X52 compounds. The base coat and the top coat are applied in a single pass through the coating and drying system. We have estimated that the OSHA AVCM concentration limit is reached with Geon 580X52 at 22mg/kg RVCM when all of the RVCM transfers to the air between the coater and the ovens, and with the ventilation rate of 6 turnovers per hour.

DISCUSSION

Introduction

In October 1974, the Occupational Safety and Health Administration (OSHA) published a safety and health standard concerning exposure of industrial workers to vinyl chloride monomer (VCM). This standard set a permissible exposure limit of 1 ppm by volume in air as a time weighted average (TWA) over an 8 hour period. Excursions up to 5 ppm are permitted over a period not to exceed 15 minutes. The standard is written for, but not limited to, employers who handle VCM. Employers in the fabricating industry are also subject to the standard where exposures approach the permissible limits. Employers handling or using fabricated products are excluded from the standard based on the absence of "adequate evidence of exposure to vinyl chloride in these operations." A fabricated product is defined as a product made wholly or partly from PVC which does not require further processing at temperatures, and at times, sufficient to cause mass melting of the PVC. In many applications latex may be considered a fabricated product and therefore the user can be excluded from the standard. Where latex may be considered as being used in the fabricating industry, exposures are not likely to approach the permissible limits.

OSHA describes an "action level" at 1/2 the permissible exposure limit, viz. 0.5 ppm TWA over an 8 hour period. Initial personal monitoring is required where vinyl chloride is used and where a product or process contains PVC are fabricated. Monitoring must be required initially where levels are measured to be >1 ppm TWA over an 8 hour period. Monitoring may be discontinued when 3 consecutive readings 5 days apart show <0.5 ppm TWA. If the employer believes that the exposure level may have increased, additional data must show 95% confidence that the exposure level is below the action level.

Table 1

0.25-
0.50-
>1.00

The following table shows the action level for vinyl chloride monomer (VCM) in air.

- 1) show a record of personal monitoring by engineering or health department, or by a qualified person, for the purpose of determining the exposure level.
- 2) keep records of personal monitoring.

- All of these requirements add expense to manufacturing operations. The expense can be avoided only by eliminating vinyl chloride or by showing that the concentration of vinyl chloride in the working environment cannot be greater than 0.5 ppm TWA for 8 hours.

The purpose of this work is to provide objective laboratory data to support what we already know: vinyl latexes stripped to 25mg/kg VCM (wet basis) can be handled, where ordinary industrial ventilation is in operation, without the likelihood of exposing workers to ambient VCM concentrations above the OSHA defined action level. With such objective laboratory data, the management of the Latex Division can support a business decision to remove cancer hazard labels from vinyl latex products. It is suspected that these labels are unnecessarily inhibiting sales of labeled latex products.

[illegible]

We have made continuous efforts since 1975 to improve stripping operations in the plants by process studies and by capital improvements. We are now able to routinely produce vinyl latex products at RVCM contents of 25mg/kg (wet basis) maximum. There is one exception, Geon 150X2, is a latex stabilized by a heat sensitive emulsifier system that cannot be stripped in the conventional manner. This product is sold to only one customer. It is acceptable to this customer at a specified RVCM content of 250mg/kg. They are able to handle the latex without OSHA compliance problems.

Plant stripping experience for most of 1980 is summarized in Table I. The average RVCM level is well below the specification level of 25mg/kg for all the major products.

TABLE I

DATA - EVCN, M2/10 - MAJOR PRODUCTS

	576	590X119	590X128	460X6 460X9	460X45 460X46	590X4	682
	7.0	6.6	7.0				
	61	60	6				
	3-10	3-10	4-9				
	10.1	10.1	10.1	5.5	9.9	3.7	12.1
	27	27	27	32	6	10	20
	4-22	4-22	4-22	9-14	3-18	0-11	0-22

III. Experimental Plan

A plan was developed which could allow us to relate AVCM concentrations to RVCN concentrations in the latex. Three latex products were chosen for this study: 1) Geon 352, our highest volume vinyl chloride resin latex, 2) Geon 576, our highest volume externally plasticized vinyl chloride resin latex, and 3) Geon 450X20, a high solids, internally plasticized, acrylic modified vinyl chloride latex. The original data set of 12 experiments included the three latexes at two latex temperatures (72 °F and 110 °F) and at two RVCN levels (as obtained from the plant and after adding a measured quantity of VCM.) The original set was later extended in efforts to improve the quality of the data correlation. A detailed description of the experimental procedure follows in the next section.

IV. Procedure - R. E. Shaffer

A. Latex Spiking

1. 15 gallons of Geon 576, Lot 200K139, were charged to a 15 gallon reactor. The reactor was sealed and the latex stirred at 100 rpm.

2. 1.5 gallons of 10% VCM solution were added to the reactor. The reactor was sealed and the latex stirred at 100 rpm.

3. 1.5 gallons of 10% VCM solution were added to the reactor. The reactor was sealed and the latex stirred at 100 rpm.

4. 1.5 gallons of 10% VCM solution were added to the reactor. The reactor was sealed and the latex stirred at 100 rpm.

5. 1.5 gallons of 10% VCM solution were added to the reactor. The reactor was sealed and the latex stirred at 100 rpm.

D. Pump Calibrations

The pumps were calibrated by personnel from the ALTC environmental group to provide an air flow rate through the charcoal tubes of ≈ 15 cc/min. The flow rate was increased to 50 cc/minute for the 15th through the 20th experimental runs where glass disposable charcoal sampling tubes from the Mine Safety Appliance Corporation were used in place of the s.e. tubes.

E. Air Sampling Pump Locations

1. For four hours preceding a latex run, duplicate "blank" air samples were taken in the closed work area. Two pumps were placed on a lab bench approximately 3.7 inches off the floor and about 1 foot from the wall.
2. During the 4 hours of a latex test, two pumps were placed side by side 37 inches off the floor and within 2 inches of the wall. Two pumps were placed in the center of the room 35 inches off the floor and directly over the latex. The two pumps (4 total) were placed 69 inches off the floor directly over the latex. Collection tubes were suspended within a few inches of the pumps. The s.e. collection tubes were pre-equilibrated before each use by placing in a beaker of water and pumping with nitrogen. The water was at room temperature and the tubes were at 25°C. The pumps were placed in the center of the room to insure that the air flow was not directly over the latex as that interference was to be eliminated.

1. After the four hours of blank samples were taken, a pail of water was placed on the floor in the center of the room.

The pumps were then used to provide good air flow in the room. After the latex was poured, latex and room temperature were recorded. Air samples were taken.

7. After 4 hours the pumps were turned off, temperatures were recorded and another 4 oz. latex sample was taken.
8. The latex samples were analyzed at ALGCP for residual vinyl chloride by gas chromatography. The analyses are routinely done there (BFG Proc. No. 1005-E).
9. The charcoal tubes were analyzed for VCM at ALTC by Pete Dysterko. The thermal exhaustion technique was used for the analysis. The analysis results were converted to ambient vinyl chloride concentration (AVCM) based upon the pumping rate, the time and the VCM adsorbed.
10. The test area was ventilated for -16 hours before collecting more "blank" samples. Latex testing conditions were, 1) spiked with VCM, 2) as received, 3) 118°F latex temperature, 4) room temperature. The air was at room temperature for all the experiments.

A schematic diagram of the experimental apparatus is shown in Fig. 1.

G. Special Test--Controlled VCM Release

1. Since the charcoal tube analyses were not as reproducible as wanted, so after run #8, a different experiment was tried. A mixture of VCM was put in the test area and a small, known amount of VCM was released continuously at a constant rate into the closed room for 4 hours. Six charcoal collection tubes and pumps were placed in the usual locations. These data are reported in the tables between tests #8 and #9. This special test was conducted by Pete Dysterko.

H. Latex Test

1. Run #11 was a latex test. A thick film over the latex started to form. At the end of the 4 hour test there was a very thick layer of solid latex surface visible. The latex was covered with foam and polymer. Later the latex was reduced to 54.0% T.S. and the surface was reduced skimming.

At the end of the test, some of the readings were taken from test #11. The readings were taken from test #11. The readings were taken from test #11.

1703777C



1. During test #7, 352 at 110°F, one of the heating elements apparently overheated and started to decompose some polymer. The test area filled with smoke. A considerable amount of smoke was detected in adjoining rooms as well. This test was stopped after 195 minutes. The tubes were later analysed as in other tests.

Glass charcoal sampling tubes from Mine Safety Appliance Corporation were used for tests #15 through #20. These replaced the stainless steel collection tubes used through test #14. Carbon disulfide extraction replaced thermal desorption as an analytical technique. The data are recorded in Table IX.

Experimental data are in books 911-83 pages 125 - 161 and 911-84 pages 1 - 18.

[illegible]

V. Testing—Air sample Tubes - Peter Dmyterko

A. Analysis for Vinyl Chloride

Standard environmental testing techniques were used to measure the concentration of vinyl chloride (VCM) in the air. Air sampling was accomplished by passing the air through charcoal collection tubes for four hours at a known flow rate. The VCM in the air was adsorbed on to the charcoal as the air was drawn through the tubes by a vacuum pump. The VCM level collected on the charcoal was then determined by one of the two following analytical procedures. In the first procedure, the VCM was thermally desorbed from the charcoal using a Sandex Flasher Unit and analyzed by flame ionization chromatography. The second procedure involved desorbing the VCM from the charcoal with carbon disulfide before gas chromatography analysis. The ppm levels of VCM in the air were then calculated from the micrograms of VCM found on a set of tubes and the volume of air sampled.

B. Sampling Procedure

[illegible][illegible]

Two tubes in series were used at each sample point. Most of the VCM was adsorbed by the charcoal in the first, and the second tube was present to collect any VCM not adsorbed by the first tube. The pump used to draw air through the tubes was a Bendix Micronair II Air Sampling Pump, a low flow rate battery operated pump.

Before sampling was started, it was necessary to calibrate through each tube set and pump used at the various sampling points. The calibration was performed by the long term six hour exposure technique described in Section VII of Procedure 1800-A. First, the charge on the pump was checked and then the pump was turned on for 15 minutes to warm it up. Next, the pump was attached to the flow rate calibration assembly together with the two conditioned collection tubes in series (see figure in Procedure 1800-A). The breathing air was turned on to the calibration assembly. The flow rate through the tubes adjusted to 7.5 to 15 cc/min. by regulating the speed of the pump with the pump flow rate screw. A soap film flowmeter and a stop watch were used to measure the exact flow rate through the tubes. After the calibration was completed, the pump was stopped and the tubes were disconnected from the calibration assembly and attached directly to the pump.

The pumps and tubes were then placed in the area to be sampled and the pumps were started. The starting time was recorded and after 6 hours of sampling, the final flow rate was checked. The pumps were stopped and the final time recorded. The flow rate was then checked. The average flow rate was calculated by dividing the total volume of air sampled by the total sampling time.

G. Analysis

The air samples were analyzed for VCM by the collection tube method. The air sample was drawn through the collection tube and the VCM was adsorbed on the charcoal. The collection tube was then placed in a desorption apparatus and the VCM was desorbed into a gas chromatograph. The gas chromatograph was calibrated with known amounts of VCM and the concentration of VCM in the air sample was determined.

In this procedure, the air samples were drawn through the collection tube and the VCM was adsorbed on the charcoal. The collection tube was then placed in a desorption apparatus and the VCM was desorbed into a gas chromatograph. The gas chromatograph was calibrated with known amounts of VCM and the concentration of VCM in the air sample was determined.

two minutes in order to desorb the VCM from the charcoal. While this step was performed, the tube was isolated from the rest of the system. Next, the flasher was placed in the "Inject" mode of operation permitting carrier gas to pass through the tube. The VCM was flash/desorbed from the charcoal and was carried on to a gas chromatography column in the chromatograph. After this column separated the VCM from other organics in the sample, the VCM was detected by a flame ionization detector. The area of the VCM peak produced was measured by an integrator. The VCM peak area was multiplied by a calibration factor to convert it to micrograms of VCM via the External Calibration Procedure. The micrograms of VCM on both tubes in a series were combined to obtain the total micrograms of VCM. Finally, the ppm levels of VCM in the air was calculated using the micrograms of VCM on the tubes and the volume of air sampled, using a formula found in Section 1.4 of EPA Method No. 821-B.

The gas chromatography conditions used in the thermal desorption procedure are listed in Table III.

Table III

Gas Chromatography Conditions for the
Thermal Description VCM Analysis

Gas Chromatograph - Hewlett and Packard (H&P)
5720A Gas Chromatography with
H&P Model 18788A Electrometer
and H&P 18791A oven control,
flame ionization detector

Integrator - H&P 3380A integrator

Column - 2 ft x 6 ft x 1/8 inch O.D. stainless
steel column packed with Porapak Q[®] 80/100
mesh (Mallory Co.)

Nitrogen
100°C
20 cc/min.

Carrier gas
100°C
20 cc/min.

Detector
100°C
20 cc/min.

Injection
100°C
20 cc/min.

Sample
100°C
20 cc/min.

Effluent
100°C
20 cc/min.

Exhaust
100°C
20 cc/min.

Reagent
100°C
20 cc/min.

Carrier gas
100°C
20 cc/min.

Detector
100°C
20 cc/min.

Injection
100°C
20 cc/min.

Sample
100°C
20 cc/min.

Effluent
100°C
20 cc/min.

Exhaust
100°C
20 cc/min.

Reagent
100°C
20 cc/min.

The 2 ft and 6 ft sections of the column
were connected in series permitting
the 2 ft section to be used for the
analysis of low boiling components
and the 6 ft section for the analysis of
high boiling components. The effluent
from the 6 ft section was also main-
tained as an independent source of carrier gas.

Analysis was performed on the H&P 3380A Integrator;
Standard Techniques

The analysis was performed at 100°C for 2.4 minutes.
The carrier gas flow rate was 20 cc/min. The analysis
was started.

The procedure for calibrating the analysis is described in Section VIII of Procedure No. 1021-A. The method involves loading conditioned collection tubes with known levels of VCM from calibration gas mixtures. To calibrate a tube, it is connected to the calibration gas mixture, and the gas mixture is allowed to flow through the tube while the VCM in the gas mixture is adsorbed on to the charcoal. The flow rate through the tube and total time the tube is exposed to the gas mixture are measured. Flow rates were typically in the range of 30 to 40 cc/min and total exposure times were in the order of 4 to 10 minutes. The micrograms of VCM adsorbed by the tube is calculated from the flow rate, exposure time and concentration of VCM in the gas mixture. The flow rate through the tubes was measured by means of a soap bubble flow meter and a stopwatch. The calibration was performed with certified gas mixtures (Matheson Co.) containing: 1.09 ppm, 10.2 ppm and 105.9 ppm VCM in air. Duplicate tubes were loaded at the following approximate levels of VCM: 1 ppm, 4 ppm and 40 ppm of VCM. The tubes were analysed for VCM using the thermal desorption procedure described above and the area for the VCM peak was obtained. A calibration factor for each tube was calculated by dividing the micrograms of VCM on the tube by the VCM peak area. An overall calibration factor was obtained by averaging the factors obtained for all the tubes loaded. The response factors were found to be linear over the calibration range.

In some of the runs, impurity peaks came close enough to VCM peak to prevent full baseline resolution of the VCM peak. In these cases, the HP Instrument employed a built-in computer program to determine the area under the VCM peak. When this situation occurred for a run, the word "interference" was indicated just after the data for the run in the data sheets.

No attempt was made to determine whether other components were present along with VCM in the VCM peak. The presence of other components under the VCM peak would cause an error on the high side for the concentration of the sample. There was no indication of this problem, however.

Occasionally, anomalous results were obtained for one set out of three sets of results which were sampling the same area over the same time period. A thorough check of the analytical system indicated that the problem was not related to chromatography analysis. The problem was never observed during the analysis of tubes during the calibration of the analysis. The cause of this problem has not been determined.

2. Carbon Disulfide Desorption Analysis For VCM

The second procedure used in the analysis of VCM adsorbed on the charcoal tubes was the carbon disulfide desorption technique. The method employed was a modification of NIOSH Method No. P&CAM 178, "Vinyl Chloride in Air" (see Appendix). The modifications in the column used and the calibration procedure were based on recommendations of the Corporate Environmental Lab at BFG Research and Development Center.

The analysis is conducted by first desorbing the VCM on the charcoal in carbon disulfide. The charcoal contents of the glass collection tubes were placed into 2 cc glass vials (Wheaton Co., 223682). Then, 1 cc of carbon disulfide (Fisher Co., 8C-182) was pipetted into each vial. The vial contents were next crimp sealed with aluminum caps containing a teflon coated rubber seal. Finally, the sample vials were agitated for 1/2 hour at room temperature on an SSC Charcoal Desorber.

Next, one microliter of the carbon disulfide phase of the vials was analyzed by flame ionization chromatography using the conditions listed in Table IV. An authentic sample of the drug in sample 10 was used as the standard. From the gas chromatogram of the sample, it was determined that the sample was 100% pure. The drug was not detected by the flame ionization detector was determined by analysis of sample 10. Enter showed 100% pure.

The first of these is the fact that the
 second of these is the fact that the
 third of these is the fact that the
 fourth of these is the fact that the
 fifth of these is the fact that the
 sixth of these is the fact that the
 seventh of these is the fact that the
 eighth of these is the fact that the
 ninth of these is the fact that the
 tenth of these is the fact that the

Table IV

Chromatography Conditions for VCM
Analysis by the CS₂ Desorption Technique

Chromatograph	=	H & P 7620A Flame ionisation chromatograph
Detector	=	Flame ionisation detector
Integrator	=	Perkin Elmer Sigma 10 Data System
Column	=	20 ft x 1/8 inch O.D. x stainless steel x 100 split on Supelcoport 80/100 mesh
Column Temperature	=	50°C, isothermal
Detector Temperature	=	250°C
Integrator Temperature	=	250°C
Injection Port Temp	=	250°C
Carrier Gas Flow	=	30 cc/min
Sample Flow	=	100 µl/min

1. *Chlorophyll a* (Chl *a*)

- 20 -

The analysis was calibrated by first loading duplicate glass charcoal tubes at the following VCM levels: 1 µgm, 4 µgm, and 40 µgm VCM. The loading procedure used was identical to that described in the calibration procedure for the thermal desorption.

tion method. The charcoal in each of these standard tubes was then desorbed in small vials with carbon disulfide as described above. Duplicate analyses were performed on each vial using the above chromatography procedure. The two VCM


peak areas obtained for each vial were averaged to obtain the peak area for the standard. A calibration factor for each standard was calculated by dividing the micrograms of VCM loaded on the tube by the average area obtained from the analysis of the standard. In general, calibration factors were

NO 3-40 10 SHEET PER SHEET
10 X 10 PER 1000

DISTENSION CORPORATION
MADE IN U.S.A.

Fig. 2 AVOM Concentration vs. Latex
AVOM Concentration--Raw Data

Note: Straight line connects the
origin and the coordinates repre-
senting the action level and the
current product specification limit



Data were submitted to the Statistics and Computer Group for analysis. Details of this analysis are shown in the Appendix. Critical RVCN levels were determined for the three latexes tested. The action level is assured with 95% confidence at 0.5 ppm AVCM if the VCM concentration in the air is 0.33 ppm. The level of 0.33 ppm, therefore, was used in the correlation to determine the critical RVCN concentration in the latex. Results are summarized in Table VI. As explained in the statistical report only 34 of the air sample test results were used; 14 were rejected because carbon tube loadings were out of the range for good accuracy; 4 were rejected because of interference with VCM in the GC analysis.

<u>Latex</u>	<u>Temperature, °F</u>	<u>Critical RVCM, mg/kg (wet basis)</u> (relates to 0.33 ppm AOC in 4 hours)
Gen 576	72	53
Gen 576	110	22
Gen 450K20	72	27
Gen 450K20	110	38
Gen 352	72	>50
Gen 352	110	>50

2) none of the above. 2) almost working environment, 3) none of the above. 3) none of the above. 4) none of the above. 5) none of the above. 6) none of the above. 7) none of the above. 8) none of the above. 9) none of the above. 10) none of the above. 11) none of the above. 12) none of the above. 13) none of the above. 14) none of the above. 15) none of the above. 16) none of the above. 17) none of the above. 18) none of the above. 19) none of the above. 20) none of the above. 21) none of the above. 22) none of the above. 23) none of the above. 24) none of the above. 25) none of the above. 26) none of the above. 27) none of the above. 28) none of the above. 29) none of the above. 30) none of the above. 31) none of the above. 32) none of the above. 33) none of the above. 34) none of the above. 35) none of the above. 36) none of the above. 37) none of the above. 38) none of the above. 39) none of the above. 40) none of the above. 41) none of the above. 42) none of the above. 43) none of the above. 44) none of the above. 45) none of the above. 46) none of the above. 47) none of the above. 48) none of the above. 49) none of the above. 50) none of the above. 51) none of the above. 52) none of the above. 53) none of the above. 54) none of the above. 55) none of the above. 56) none of the above. 57) none of the above. 58) none of the above. 59) none of the above. 60) none of the above. 61) none of the above. 62) none of the above. 63) none of the above. 64) none of the above. 65) none of the above. 66) none of the above. 67) none of the above. 68) none of the above. 69) none of the above. 70) none of the above. 71) none of the above. 72) none of the above. 73) none of the above. 74) none of the above. 75) none of the above. 76) none of the above. 77) none of the above. 78) none of the above. 79) none of the above. 80) none of the above. 81) none of the above. 82) none of the above. 83) none of the above. 84) none of the above. 85) none of the above. 86) none of the above. 87) none of the above. 88) none of the above. 89) none of the above. 90) none of the above. 91) none of the above. 92) none of the above. 93) none of the above. 94) none of the above. 95) none of the above. 96) none of the above. 97) none of the above. 98) none of the above. 99) none of the above. 100) none of the above.

[illegible]

VII. Translation of Critical VCM Data to Other Vinyl Latex Products

When the lid is removed from a container of vinyl latex, VCM concentration gradients are created. This causes diffusion of the monomer to occur. There are several steps in series that the monomer must take in the diffusion process. Picture latex in a closed container. The system is in static equilibrium. The monomer is distributed between the phases: polymer, water, and vapor. The concentration in each of the phases depends upon the solubility of vinyl chloride in the polymer and in the water.

Equilibrium data are available for six vinyl latex products representative of the three types: resin, externally plasticized, and internally plasticized. The data relate vapor concentrations in a closed system with concentrations in the latex over a temperature range. An attempt has been made to determine the distribution of monomer between the water phase and the polymer phase. It has not been important for our purposes to do this. The equilibrium data have been fitted to the mathematical model of this form:

$$P_{VCM} = K_1 W^2 e^{-K_2/T}$$

where: P_{VCM} = partial pressure of VCM, mm Hg
 W = weight percent VCM in the latex at 57° F.
 T = absolute temperature, °K
 K_1 , K_2 and A are correlation constants

Values for the correlation constants are shown in Table VII.

TABLE VII
 Correlation Constants for Vapor-Liquid Equilibria

Product	Wt % VCM	T, °K	K_1	K_2	A
			24.4	5413	1.00
			21.5	4741	1.06
			13.1	2230	1.03
	351		16.0	3075	1.02
576	78	351	11.03	1605	0.92
580K119	65	352	17.3	3453	0.95

The diffusion process starts when the vapor space is suddenly increased to include the air volume surrounding the opened container. The monomer vapor then begins to migrate into the new space as the system tries to establish a new equilibrium, so the monomer begins to move through these steps: 1) diffusion through the air, 2) diffusion through the water-air interface, 3) diffusion through the water, 4) diffusion through the polymer-water interface, and 5) diffusion through the polymer. It is impossible to guess which step controls the rate of increase of AVCM concentration. As a worst case, however, it can be assumed that diffusion through the water-air interface is controlling. The driving force for the diffusion then is the partial pressure of VCM at the liquid-air interface, i.e., the equilibrium partial pressure at the temperature and AVCM concentration of the latex. Equilibrium partial pressures have been calculated for the conditions shown in Table VI. The final result is the average for the six conditions: critical $p_{\text{vcm}} = 1.36 \text{ mm Hg}$. The standard deviation for this result is: $\pm 0.65 \text{ mm}$. This is a large error. The large standard deviation is not surprising considering the large error in the AVCM data. The critical AVCM concentration, 0.15 g/g , was calculated to the critical $p_{\text{vcm}} = 1.36 \text{ mm Hg}$. This pressure is used to calculate the critical AVCM for the other three polymers, shown in Table VII. For which equilibrium data are available. However, for other unsaturated products based on CMC, the critical AVCM is not available. An interpretation of critical AVCM for these polymers is not possible. The results of these critical AVCM calculations are shown in Table VIII. The allowable AVCM for these polymers was calculated the effect of resin dilution with water.

Table VIII

Calculated Critical RVCM Levels in Vinyl Latexes
Based on Equilibrium Partial Pressure

Critical Pvcm = 1.38 mm Hg, $\sigma = \pm 0.65$ mm Hg

Corresponds with AVCM = 0.33 ppm in closed, unventilated room

Latex	Plasticizer pphr	Base Resin	Calculated Critical RVCM mg/kg (110°F) (1)	Allowable RVCM Limit for Base Resin mg/kg
352	-	-	35 (50)	-
351	-	-	29	-
450X20	-	-	38 (38)	-
500X29	22.4	351	29	35
576	38	351	22 (22)	30
500X19	45	352	17	28
500X17	30	352	27 (2)	35
500X15	18	352	28 (2)	34

(1) The value in parentheses is the result from statistical analysis of the data (see Table VI)

The values were by interpolation of calculated data

It is noted from the data that for every product the critical RVCM, or the level of RVCM for base resin used to make the plasticized product, is less than the level allowed by the product specification. In some cases, the plastic products have critical RVCM levels that are less than the product specifications. With the base resin used in the product, base 352, at 28 mg/kg RVCM, from 500X19, the critical RVCM would be 18 mg/kg. The product specification for these two products is 15 mg/kg maximum and 13 mg/kg maximum.

VIII. Diffusion Rate Estimate.

It is of interest to know the rate of diffusion of VCM from latex to surrounding air space to which it is exposed. From such information it is possible to estimate possible personnel exposure levels in a working environment under known ventilation conditions. Each of our closed room experiments could provide such information. The amount of VCM transferred during the test can be calculated from latex KVCN data at the beginning and at the end of the experiment. This quantity should be in agreement with the quantity measured as the AVCM concentration during the test.

The calculations have been made based on the following assumptions:

- 1) the rate of transfer from the latex to the surrounding air is constant during the test period,
- 2) the concentration of AVCM changes linearly with time from 0 at the beginning of the test to a maximum at the end of the test,
- 3) the measured AVCM for a given test is the average concentration for the test period.

For the measured AVCM value, the average of the statistically validated air samples taken during a test period was used. Air movement within the room during a test was considered sufficient to eliminate gradients in the air space. Statistical analysis showed no correlation of AVCM concentration with position of the air sample. Results of the calculations are shown in Table I. The results are shown graphically in Fig. 3. The graph shows the AVCM concentration for the special test, described earlier, which was conducted in a room that contained a cylinder of latex at a constant level of 1.0 g/l. During this time the air in the room was maintained at 1.0 g/l in all the other tests.

The amount of VCM released during a test is calculated by air sampling.

Table IX
RVCN Calculated from RVCN Loss Data

Latex	Latex	Latex	Test	RVCN	RVCN
Sample	Loss, %	Sample	Time, min.	Loss, mg/kg	Para., mg/kg
576	56.9	576	240	19.32	55.8
576	56.9	576	240	21.46	71.3
576	56.9	576	240	3.00	14.1
576	56.9	576	240	0.80	16.1
352	56.9	352	240	16.00	50.8
352	56.9	352	240	4.00	56.8
352	56.9	352	120	-1.10	4.3
352	56.9	352	240	9.30	4.2
450X20	56.9	450X20	240	11.90	47.3
450X20	56.9	450X20	240	25.10	48.4
450X20	56.9	450X20	240	1.50	4.7
450X20	56.9	450X20	240	1.10	4.0
352	56.9	352	240	-	-
450X20	56.9	450X20	240	0.23	3.2
450X20	56.9	450X20	240	3.07	31.8
450X20	54.0	450X20	240	2.45	19.4
450X20	54.0	450X20	240	3.37	19.7
450X20	54.0	450X20	240	13.71	67.9
450X20	54.0	450X20	240	22.35	67.3
450X20	54.0	450X20	240	5.00	15.8
576	56.9	576	240	3.62	11.5

Consolidated RVCN data are excluded, number in parenthesis indicates the number of samples for the test.

THE UNIVERSITY OF CHICAGO

Fig. 1 Measured AVCH Concentration as Related to AVCH Calculated from Latex RVCH Loss During the Test

This finding is based on the overall average for all the validated tests and is represented by the dotted line in Fig. 3. The failure to detect 77% of the VCM released to the atmosphere is explained by diffusion of monomer through the porous ceiling in the closed room and through joints in wall panels. Adsorption of monomer onto surfaces in the room could also account for some of the loss. The loss is equivalent to expanding the original volume of 1000 ft³ to 4350 ft³ with new air during the 4 hour test period.

So, in a case where 0.33 ppm AVCM is the measured concentration as determined by collection of air samples over a 4 hour test period, it can be determined that 42 lb. of latex in the 5 gallon container would have to release 10.57 mg of VCM per kg of latex. And assuming that the rate of release is constant, the diffusion rate is: (11.09) (10⁻⁵) lb/hr. To generalize the result, the latex-air interfacial area (0.6303 ft²) is taken into account and the diffusion rate becomes: (1.76) (10⁻⁴) lb/ft²-hr. This then is the diffusion rate of VCM from latex to air where temperature and concentration conditions are such that the equilibrium partial pressure of VCM is 1.38 mm Hg.

Knowing the rate of evolution of VCM from the latex to the room, an equivalent air flow can be calculated that will maintain 0.33 ppm AVCM. The corresponding air flow is 2117 ft³/hr. Since the room volume is 1000 ft³, this is equivalent to 2.1 air turnovers/hr. Barker and his co-workers reported a ventilation rate of 0.4 - 1.3 air turnovers/hr. in the same type of room under similar conditions. The system under test is better ventilated than the room diffusing from dense latex sample bottles.

It is safe to assume that the diffusion rate varies linearly with the equilibrium partial pressure of the latex. To check this assumption, equilibrium partial pressures have been calculated for the three latex samples at the AVCM concentrations and latex temperatures shown in Table 1. Fig. 4 shows the statistical relationship between the partial pressures and the diffusion rates. The correlation coefficient is 0.99, which is excellent considering the poor

MS 540-10 SYSTEMS, INC.
10 N 10th Ave
New York, N.Y.

OVERSEAS CORPORATION
New York, N.Y.

Fig. 4 AVCH vs. Equilibrium
VCM Partial Pressure

Note: Dotted line goes
through the critical
coordinates: 0.33 ppm AVCH,
1.38 mm Hg.



IX. VCN Estimate for a Typical Compounding Operation

A hypothetical compounding operation has been conceived in order to calculate possible VCN exposure levels in a real work area.

Work space: volume 20 ft x 40 ft x 25 ft
(= 20,000 ft³)
air temperature 90°F
barometric pressure 750 mm Hg
ventilation variable

Compound mix tank:

volume 1000 gal.
diameter 60 in.
(liquid-air interface 2830 in²)
agitated

Notes:

Comp 576, 576 T.S., VCN variable
temperature 110°F

The air in the tank is assumed to be at equilibrium with the compound.

The liquid level in the tank is assumed to be 10 in. below the top of the tank.

The air in the tank is assumed to be at equilibrium with the compound.

The ventilation requirement is calculated to maintain the VCN concentration at 100 ppm.

The ventilation requirement is calculated to maintain the VCN concentration at 100 ppm.

The calculations are summarized in Fig. 3. Typical industrial ventilation gives 6 turnovers of air per hour. At this level of ventilation, the Geon 576 used in a typical compounding operation could be at 40 mg/kg RVCM without exceeding the OSHA action level. At the proposed specification limit for Geon 576, 18 mg/kg, only 3 air turnovers per hour are required to stay in the OSHA compliance region.

X. Paper Coating Operation at Rhineland Paper Company

Geon 576 is used at Rhineland Paper Company to apply a clear coating to paper.

Work space: volume 20 ft x 150 ft x 100 ft
(= 300,000 ft³)

air temperature 90°F

barometric pressure 750 mm Hg

ventilation 6 turnovers/hr.

Coating operation:

Coating applied 40 lb (wet)/room
(2 room x 2000 ft³)

Coating rate 100 lb/min.

Coating thickness 1/2 in.

Coating distance 10 ft.

Geon 576 2.5% w/w in 20 mg/kg

These calculations show that the use of Geon 576 in a paper coating operation could result in a concentration of 2.5% w/w in the air, which is well below the OSHA action level of 10 mg/kg. The use of Geon 576 in a paper coating operation is a typical industrial application of this material. The use of Geon 576 in a paper coating operation is a typical industrial application of this material. The use of Geon 576 in a paper coating operation is a typical industrial application of this material.

MS-800-10 Electronic Graph Paper
10 x 11, 425 x 425

GREENSBORO, N.C.
MAY 1964

Fig. 5 Ventilation
Requirement for NSHA
Compliance



The working environment at Shinslander was monitored during a coating operation using Gen 376. Results were reported by two different laboratories accredited by AIAA:

1) the BPGoodrich Corporate Environmental Health Laboratory in Brecksville, Ohio, and

2) Wausau Insurance Companies, Wausau, Wisconsin.

These reports are included in the Appendix of this report. Results are consistent with our calculated data.

II. Serim Coating Operation at Sanitas

Gen 36X33 is used as a pigment binder at Sanitas in a double coating operation. A base coat compound is applied to the scrim on the first roller. The base coat is dried in an oven. A top coat compound, also based on 36X33, is then applied and the material is again dried in an oven. All these operations take place in one pass through the coating line.

Gen 36X33 is used as a pigment binder at Sanitas in a double coating operation.

Again, the worst case situation is assumed, i.e., the unlikely situation where all of the VCM in the two compounds would transfer to the surrounding work space. The latex is consumed at a rate of 1134 lb/hr in the base coater and at 303 lb/hr in the top coater. The total consumption rate is 1437 lb/hr and the VCM evolved is 0.0359 lb/hr.

$$P_{vcm} = \frac{VCM}{NV} = \frac{(0.0359) (355) (550)}{(62.4) (2.4) (10^6) (6)} = 1.22 (10^{-5}) \text{ mm Hg}$$

And the AVCM concentration is

$$AVCM = \frac{(1.22) (10^{-5}) (10^6)}{750} = 0.016 \text{ ppm}$$

This is well below the OSHA action level. There is no likelihood that the use of these 550 lb compounds at 11 minutes will create an OSHA compliance problem.

CHODNYEC

The subjects are grateful for help from the Environmental Control
Agency, E. B. Harwood, R. E. Hutton, and E. V. Hittig who
provided the necessary equipment and supplies; and for data
collection of specimens in the system group; Paul Cline,
and the "Biology" group, which includes, and has been
conducted by E. B. Harwood, R. E. Hutton, and E. V. Hittig of the
Agency.

ACKNOWLEDGMENTS

LIST OF REFERENCES

1. Data Book 911-63-128 through 161.
2. Data Book 911-66-1 through 10.
3. Satava, E.D., Interorganization Correspondence,
Re: VCM Label Removal, NDS-125-60, 11-18-60.
4. Osterud, E.B., Letter to Customers, Re: Green Label-Removal
of Many Chloride Warning Labels, 11-19-60.
5. Parker, G.E. and Walker, E.G., SRS Technical Document,
Removal of Green-Hazard Labels from Green Labels,
Manufacture and Computer Applications Report, 12-9-60.

6. Parker, G.E. and Walker, E.G., SRS Technical Document, Removal of
Green-Hazard Labels from Green Labels, Manufacture and Computer Applications Report, 12-9-60.

7. Parker, G.E. and Walker, E.G., SRS Technical Document, Removal of
Green-Hazard Labels from Green Labels, Manufacture and Computer Applications Report, 12-9-60.

BFG TECHNICAL DOCUMENT

Statistics and
THE B F GOODRICH COMPANY CHEMICAL DIVISION Computer Applications REPORT

Author D.S. Farley-D.S. Miller	Category ALTC	Date 12/9/80	Can Book Reference
Reviewed By [illegible]	Category ALTC	Project No. 3094-61	Shipment Date

REMOVAL OF "CANCER-HAZARD" LABELS FROM GRN LATEXES

BACKGROUND:

Marketing wants to remove the "cancer-hazard" labels from GRN latexes containing residual vinyl chloride monomer (VCM). BFG was asked to demonstrate that concentrations of VCM diffusing out of stripped GRN latexes are below the 0.5 ppm action level defined by OSHA.

Removal of "Gaseous-Solvent" Labels from OSHA Labels

12/3/80

Page 2

DATA

The original data set consisted of 12 experiments and 72 samples of airborne vinyl chloride monomer (AVCM). Only 54 AVCM samples were usable. Analytical data for fourteen samples were rejected because the VCM loadings (mg VCM/activated carbon sampling tube) were outside of the 1.0 mg to 10.0 mg range recommended by P. H. Kahrst and J. W. Horn (NIEC). Data on four samples were not usable because of interference during the gas chromatographic analysis.

The experimental conditions used are summarized in Table 1, below. A complete list of data used in the correlation analysis can be found in the Appendix (Table 1-4). Rejected data are listed in Table 2-4 of the Appendix.

TABLE 1

Summary of Experimental Conditions

Expt. No.	Airflow Rate (L/min)	Label Type	Flow Rate for Sampling Tube (L/min)	Analytical Method
1	10	1	10	Gas Chromatography
2	10	1	10	Gas Chromatography
3	10	1	10	Gas Chromatography
4	10	1	10	Gas Chromatography
5	10	1	10	Gas Chromatography
6	10	1	10	Gas Chromatography
7	10	1	10	Gas Chromatography
8	10	1	10	Gas Chromatography
9	10	1	10	Gas Chromatography
10	10	1	10	Gas Chromatography
11	10	1	10	Gas Chromatography
12	10	1	10	Gas Chromatography
13	10	1	10	Gas Chromatography
14	10	1	10	Gas Chromatography
15	10	1	10	Gas Chromatography
16	10	1	10	Gas Chromatography
17	10	1	10	Gas Chromatography
18	10	1	10	Gas Chromatography
19	10	1	10	Gas Chromatography
20	10	1	10	Gas Chromatography
21	10	1	10	Gas Chromatography
22	10	1	10	Gas Chromatography
23	10	1	10	Gas Chromatography
24	10	1	10	Gas Chromatography
25	10	1	10	Gas Chromatography
26	10	1	10	Gas Chromatography
27	10	1	10	Gas Chromatography
28	10	1	10	Gas Chromatography
29	10	1	10	Gas Chromatography
30	10	1	10	Gas Chromatography
31	10	1	10	Gas Chromatography
32	10	1	10	Gas Chromatography
33	10	1	10	Gas Chromatography
34	10	1	10	Gas Chromatography
35	10	1	10	Gas Chromatography
36	10	1	10	Gas Chromatography
37	10	1	10	Gas Chromatography
38	10	1	10	Gas Chromatography
39	10	1	10	Gas Chromatography
40	10	1	10	Gas Chromatography
41	10	1	10	Gas Chromatography
42	10	1	10	Gas Chromatography
43	10	1	10	Gas Chromatography
44	10	1	10	Gas Chromatography
45	10	1	10	Gas Chromatography
46	10	1	10	Gas Chromatography
47	10	1	10	Gas Chromatography
48	10	1	10	Gas Chromatography
49	10	1	10	Gas Chromatography
50	10	1	10	Gas Chromatography
51	10	1	10	Gas Chromatography
52	10	1	10	Gas Chromatography
53	10	1	10	Gas Chromatography
54	10	1	10	Gas Chromatography

The experiment was conducted inside a closed, unventilated room. The room was equipped with a fan and a pump to sample the air in the room.

Removal of "Cancer-Related" Labels from OSHA Labels

12/9/80

Page 3

DISCUSSION (Cont'd)

A. OSHA Regulations (continued)

If NIOS could demonstrate that the NVCH levels in OSHA labels will not produce airborne VCH levels in excess of the 0.5 ppm action level, the "cancer-related" labels can be removed from these products. Users of NIOS vinyl labels would not need to establish programs for monitoring employee exposures.

B. Simulating OSHA Test Procedures

The procedure OSHA uses to determine employee exposure to airborne VCH is outlined in Figure 1. A program was developed which simulates this procedure. The program calculates the level of airborne VCH and a specific level of NVCH associated with VCH detection and monitoring. The program also calculates the monitoring and testing error, based on NVCH levels. The program is described in Figure 2.

The program also calculates the monitoring and testing error, based on NVCH levels. The program is described in Figure 2.

The program also calculates the monitoring and testing error, based on NVCH levels. The program is described in Figure 2.

The program also calculates the monitoring and testing error, based on NVCH levels. The program is described in Figure 2.

Report of Research on the Effect of Viny1 Latexes

12/9/70
Page 6

DISCUSSION (Cont'd)

C. In Viny1 Latexes (continued)

All these latexes tested will comply at the 99% level to OSHA regulations, if they are stripped to the levels proposed in the following Table 2.


TABLE II

Critical Levels of NVCM in Viny1 Latexes

Room Temp.	Critical NVCM level (ppm by wt.)	Proposed NVCM level (ppm by wt.)
20	20	20
25	25	25
30	30	30
35	35	35
40	40	40
45	45	45
50	50	50
55	55	55
60	60	60
65	65	65
70	70	70
75	75	75
80	80	80
85	85	85
90	90	90
95	95	95
100	100	100

Removal of "Tobacco-Related" Labels from Some Licenses
1/7/50
Page 5

FIGURE 1
DATA
SUPPLYING PROBABLY



Removal of "Giant-Killer" Labels from Open Labeled

11/9/80

Page 6

FIGURE 3



FIGURE 4



Page 2 of 2

Summary of Testable AFCE Data

[illegible]

Removal of "Concentrator" Labels from CHN Latexes
 11/4/60
 Page 8

APPENDIX (Cont'd)

Table 2-4

Rejected Data

Latex	Average KVM in Latex (mm by vol.)	Latex Temp. (°F)	KVM Level (mm by vol.)	KVM Loading (mm)
CHN 352	6.3	110°	0.03	0.327
"	"	"	0.02	0.133
"	"	"	0.04	0.233
"	"	"	0.04	0.244
"	"	"	0.05	0.312
CHN 377	26.8	72°	0.09	0.730
"	2.2	72°	0.02	0.207
"	"	"	1.00	13.90
"	20.0	120°	.20	.25
"	"	"	.07	.021
"	22.05	120°	1.00	13.02
"	"	"	.04	0.377
"	"	"	1.12	17.34
"	"	"	1.27	17.44
(Superfrenon)				

[illegible][illegible]

TITLE: SAMPLING PROCEDURES FOR VINYL CHLORIDE IN AIR USING THE DOWTHERM PERSONNEL MONITORING COLLECTION METHOD	TYPE: ORDINARY Absorption Tube	SUBJECT: INDUSTRIAL HYGIENE AIR AND RELATED AIR SAMPLES
--	-----------------------------------	--

NOISFLA TYCHIC NOYCOOCH

III. INTERFERENCE

1. Flashing and analysis of each tube does not always remove 100% of the VCM from the previous collection period. It is therefore necessary to "pre-flash" or condition each FWCC in the Bendix Manifold Flasher to remove residual material which may remain from the previous analysis.
2. After conditioning and prior to sample collection, the FWCC's are to be capped and stored in a desiccator. Maximum storage time is 24 hrs. Moisture pick-up will reduce the capacity of the FWCC.
3. After the exposure period and prior to laboratory analysis, the FWCC's are to be stored in a desiccator or freezer. A separate desiccator is required for storage of exposed tubes. It is recommended that the maximum aging time should not exceed 24 hours.
4. High sample air flow rates for extended time periods contribute to "breakthrough" (leak of VCM during sampling). Calibrate the flow rate to the prescribed "safe" flow rates as indicated in this manual. Flow rate tubes are also used to indicate if breakthrough occurs.

High sample air flow rates contribute to breakthrough in the FWCC's. High flow rates also contribute to VCM on the FWCC and thereby reduce the capacity of the FWCC. It is therefore necessary to calibrate the flow rate to the prescribed "safe" flow rates as indicated in this manual. Flow rate tubes are also used to indicate if breakthrough occurs.

High sample air flow rates contribute to breakthrough in the FWCC's. High flow rates also contribute to VCM on the FWCC and thereby reduce the capacity of the FWCC. It is therefore necessary to calibrate the flow rate to the prescribed "safe" flow rates as indicated in this manual. Flow rate tubes are also used to indicate if breakthrough occurs.

IV. REMARKS

IV. SAFETY (cont'd.)

2. Observe all normally accepted safe operating practices pertaining to the specific area being sampled.
3. Use appropriate safe handling procedures associated with compressed air cylinders.
4. Observe all normally accepted safe practices in handling electrically heated equipment.

V. APPARATUS

A. Sampling Equipment

1. Model 1000-11 - Portable air sampling pump, Part No. 1000-11, equipped with battery recharger, Part No. 1000-12, or suitable unit recharger - Part No. 1000-13.

2. Model 1000-14 - Portable air sampling pump, Part No. 1000-14, equipped with battery recharger, Part No. 1000-12, or suitable unit recharger - Part No. 1000-13.

3. Model 1000-15 - Portable air sampling pump, Part No. 1000-15, equipped with battery recharger, Part No. 1000-12, or suitable unit recharger - Part No. 1000-13.

4. Model 1000-16 - Portable air sampling pump, Part No. 1000-16, equipped with battery recharger, Part No. 1000-12, or suitable unit recharger - Part No. 1000-13.

5. Model 1000-17 - Portable air sampling pump, Part No. 1000-17, equipped with battery recharger, Part No. 1000-12, or suitable unit recharger - Part No. 1000-13.

6. Model 1000-18 - Portable air sampling pump, Part No. 1000-18, equipped with battery recharger, Part No. 1000-12, or suitable unit recharger - Part No. 1000-13.

V. APPARATUS (cont'd.)

B. Calibration Assembly (cont'd.)

4. EICO, Battery Tester - Model 584. EICO Electronics, Flushing, N.Y. or equivalent.
5. Stainless steel tubing 1/4" and the appropriate fittings to be installed between breathing air source and the PMOC (see attached diagram of the calibration assembly).
6. Tygon tubing 1/4" I.D. and the appropriate tubing connectors.
7. Squeegee
8. Stainless steel air reservoir with vent holes. Approximately 200 cc volume. Fabricated as per diagram.

C. Metallurgical Equipment

1. Sling Psychrometer, Bacharach Inst. Co., Pittsburgh, Pa., +20 to 120°F, Part No. 20-511 or equivalent.
2. Two stainless steel mandrels. One are required, one for exposed surface and one for finished tubes.

1. The purpose of this step is to remove any residual moisture from the tubing. This process may be repeated if necessary at intervals as indicated in the diagram.

2. The purpose of this step is to remove any residual moisture from the tubing. The use of a squeegee is recommended.

3. The purpose of this step is to remove any residual moisture from the tubing. The use of a squeegee is recommended.

4. The purpose of this step is to remove any residual moisture from the tubing. The use of a squeegee is recommended.

5. The purpose of this step is to remove any residual moisture from the tubing. The use of a squeegee is recommended.

6. The purpose of this step is to remove any residual moisture from the tubing. The use of a squeegee is recommended.

7. The purpose of this step is to remove any residual moisture from the tubing. The use of a squeegee is recommended.

VI. CONDITIONING PMCC'S FOR USE (cont'd.)

3. It is recommended that PMCC's stored more than 24 hours be reconditioned prior to use. All new PMCC's must be conditioned prior to initial exposure.

VII. BENDIX PUMP CALIBRATION

For the purposes of both short term "task" sampling breathing air analyses as well as long term "time weighted average" breathing air sampling, there are two standardized sampling time periods associated with specific pump flow rates. The following pump calibration procedure covers the long term, 6-hour exposure period. For instructions on short term sampling, refer directly to step 10 of this section.

1. Remove the Bendix Micronair II pump from its battery charger and record its identification number along with the identification numbers of the primary and secondary conditioned PMCC's in the appropriate space on the laboratory data form.

2. Check the pump battery with the battery tester. Do not use if not fully charged.

3. Turn the pump on and let it run on the bench for 15 minutes while the battery warms up.

4. Connect the calibration assembly along with the two PMCC's to the pump. The outlet end of the PMCC is to be connected to the breathing air supply source. The Bendix pump is used to operate the PMCC's. See the instructions of the flow rate calibration assembly.

5. Connect the breathing air supply to a flow rate of approximately 1.0 cfm.

6. Adjust the pump flow rate screw to achieve a flow rate of 1.0 cfm. The height of the water column on the U-tube manometer should be approximately 2 inches.

7. The flow rate of the pump is checked approximately every 1.0 cfm. The flow rate is checked by the following procedure:

8. The flow rate is checked by the appropriate space on the laboratory data form. The flow rate is checked by the appropriate space on the laboratory data form. The flow rate is checked by the appropriate space on the laboratory data form.

6. Turn off the pump and disconnect the PMCC's and pump from the flow rate calibration assembly. Connect the tandem PMCC's to the pump using a length of Tygon tubing and an adapter. The primary PMCC is the one farthest from the pump. Again the scrub end of each tube is to be directed away from the pump.

9. Repeat calibration steps 1 thru 8 for each pump - PACE sample collector combination. Pump - sample collection combinations as calibrated must be used within 8 hours. If not exposed within this period, pumps are to be put back on charge and PACE's are to be reconditioned. Recalibration is also necessary prior to use.

10. There are two basic sample collection time intervals, 6 hours, and 15 minutes. The permissible time exposure range for the 6 hour sampling condition is 4-7 hours, while that of the 15 minute sampling condition may range from 15 to 30 minutes. The above collection conditions are for the 6 hour exposure periods. The collection procedure is similar, except for the 15 minute exposure period, that the following conditions, for exposure times of 15 to 30 minutes, shall be used. Flow and duration to be 10/min. and 15 to 30 minutes, respectively. The flow, for 15 to 30 minutes, shall be 10/min. and the duration shall be 15 to 30 minutes.

Wang.

... and personal identification

...to enter the sample collection unit
...period, and to perform his daily
...arrangements and to be made at this
...interval at the end of the sampling
...of designated job duties and locations.
...month's "work" sampling of high WCM

ALL INFORMATION CONTAINED HEREIN IS UNCLASSIFIED DATE 08-19-2010 BY 60322 UCBAW

VIII. SAMPLING PROCEDURE (cont'd.)

7. Note should be made of respiratory symptoms by the employee during the sample collection period.
8. At the conclusion of the sample collection, first, the pump flow rate is to be measured. The flow rate is to be measured in the field using a portable flow meter and stopwatch. The pump is then turned off and the final flow rate in cc/min. and the time when the pump was turned off are recorded.

IX. REPORTING AND CALCULATIONS

1. Report both the initial and final flow rates (see Appendix B2). Average these to obtain the cc/minute average flow rate over the exposure period. Calculate and record the total volume of air breathed in c.c.'s by multiplying the average cc/minute flow rate by the total minutes of exposure. Record this value on the data sheet.

Record and report the ambient temperature and relative humidity at the time and place where the sample was collected. Report any other pertinent information.

Report any symptoms of the subject concerning

any respiratory distress, cough, or other symptoms. Report any other pertinent information. Report any other pertinent information.

X. APPENDIX (cont'd.)

2. Other than battery replacement, there is little maintenance which can be performed on the pumps themselves. However, it is recommended that pump & fall-off rate records be kept for each pump. As much as 15% fall-off in flow rate from the start to the end of an exposure period can be considered normal. Percent fall-off rates greater than this should be investigated for battery failure or pump problems. (Note: This is considered normal for the Bendix, Micronair II pumps. Sigis or other NIOSH approved pumps may exhibit different fall-off flow rate characteristics.)
3. To prevent coupling errors and false breakthrough, the O-rings in the tube holder and back-up conversion kits (tandem tube adapters) must be replaced every six months. Replace them with 1/4" I.D., 1/8" O.D., Viton O-rings. Use Viton O-rings when they are exposed to hazardous chemical service.

SHIFT _____

PTMAGNET MONITORING - VINYL CHLORIDE MONITOR, BONDEN FLASHER


DATE 1

Location: _____

Page 01

1	2	3	4	5	6	7	8	9
TIME	DATE	VINYL CHLORIDE MONITOR	MONITOR	MONITOR	MONITOR	MONITOR	MONITOR	MONITOR
		YES/NO	YES/NO	YES/NO	YES/NO	YES/NO	YES/NO	YES/NO
		YES	YES	YES	YES	YES	YES	YES
		NO	NO	NO	NO	NO	NO	NO
		YES	YES	YES	YES	YES	YES	YES
		NO	NO	NO	NO	NO	NO	NO
		YES	YES	YES	YES	YES	YES	YES
		NO	NO	NO	NO	NO	NO	NO
		YES	YES	YES	YES	YES	YES	YES
		NO	NO	NO	NO	NO	NO	NO
		YES	YES	YES	YES	YES	YES	YES
		NO	NO	NO	NO	NO	NO	NO
		YES	YES	YES	YES	YES	YES	YES
		NO	NO	NO	NO	NO	NO	NO
		YES	YES	YES	YES	YES	YES	YES
		NO	NO	NO	NO	NO	NO	NO

BENDIX PUMP CALIBRATION ASSEMBLY



R.F. GOODRICH CHEMICAL COMPANY
STANDARD TEST PROCEDURE No. 1021-A

Page 1 of 1

TITLE: Analysis of Vinyl Chloride Monomer in Air Via Carbon Collection Tube Thermal Desorption

TYPE: Gas Chromatograph

PROJECT: Ambient Breathing Air and Related Air Samples

* Sample collection method to be covered by separate procedure (see Note 3)

I. SCOPE

The use of the thermal desorption technique incorporated in this procedure is suitable for monitoring the level of vinyl chloride vapors in air in relation to personnel exposures. The method, with modifications, can also be utilized for determining exposure to other organic vapors.

II. SUMMARY

This system of sampling and analyzing industrial vapor enables the collection tubes to be connected directly to the G.C. analytical column after sample collection. The organic vapors are thoroughly flushed/desorbed from the carbon based collection tube into a laboratory gas chromatograph. The results are obtained in the form of vinyl chloride absorbed during the sampling period. A column backflushing technique is utilized to remove contaminants of high boiling components.

III. LIMITATIONS

Interfering compounds which elute from the chromatographic column at the same time as vinyl chloride will adversely affect the accuracy of this analysis.

Collection tubes, either samples or standards, beyond the recommended shelf life should not be used. Refrigeration (-15°F) of exposed samples for a maximum tube storage of 24 hours without appreciable VCM degradation. Collection tube standards can be frozen along with samples to prevent degradation during storage.

Standardize against known standards, and to insure use of the Century equipment.

		ANALYST		DATE
Sample Name	10/1/74	J. E. Lab	J. E. Gammali	11/15/74
Sample No.	10/1/74	10/1/74	10/1/74	
Sample No.	10/1/74	10/1/74	10/1/74	
Sample No.	10/1/74	10/1/74	10/1/74	
Sample No.	10/1/74	10/1/74	10/1/74	
Sample No.	10/1/74	10/1/74	10/1/74	
Sample No.	10/1/74	10/1/74	10/1/74	
Sample No.	10/1/74	10/1/74	10/1/74	
Sample No.	10/1/74	10/1/74	10/1/74	
Sample No.	10/1/74	10/1/74	10/1/74	

III. INTERFERENCES (cont'd.)

3. All collection tubes must be pre-flashed prior to their initial use, and must be conditioned prior to each subsequent exposure.

IV. PRECISION

An accurate determination of the total test reproducibility of this method has not as yet been defined. But the standard deviation of the ability to reproduce on calibration standards has been measured at approximately 7% at all three calibration levels of 0.5 μ g, 45.0 μ g and 450 μ g. Data on calibration standards was collected daily over a one month period.

V. SAFETY

1. Vinyl chloride monomer is a suspected carcinogen. Do not release vinyl chloride vapors to the laboratory atmosphere during standard preparation. Standard preparation should be performed in a suitable, properly functioning, lab fume hood and the vapor must be vented to the outside air.
2. Do not allow the tube to come into contact with heated parts of the collection tube rack or desiccator. Flammable/combustible tubes are to be removed from the rack with the needle-valve closed and should be allowed to cool before the valve is opened.
3. Do not use the tubes for any other purpose than that for which they were designed.
4. Do not use the tubes for any other purpose than that for which they were designed. Do not use the tubes for any other purpose than that for which they were designed. Do not use the tubes for any other purpose than that for which they were designed.
5. Do not use the tubes for any other purpose than that for which they were designed. Do not use the tubes for any other purpose than that for which they were designed. Do not use the tubes for any other purpose than that for which they were designed.

VI. APPARATUS (cont'd.)

4. 2 ft. and 4 ft. x 1/8" stainless steel columns packed with Poropak Q 30/60 mesh (Maters Associates, Inc.), or equivalent substitute.
5. Standard cylinders, vinyl chloride in air at 10, 100, and 1000 ppm by volume, Matheson Gas Products, Joliet, Illinois. When ordering, specify zero grade air (vinyl chloride in nitrogen can be used instead of air if desired).
6. Pressure regulators, Fo. EL (GCA) Matheson Gas Products (3) required for item 5 above.
7. Reducator, No. 7221T, 65 mm, aluminum, 1/4" tubing, Matheson Gas Products, Joliet, Illinois. (To be installed downstream of the pressure regulator in each of the calibration gas lines.)
8. Hydrogen cylinder with regulator.
9. Nitrogen cylinder with regulator.
10. Air cylinder with regulator.
11. Emergency air supply and regulator.
12. Safety valve assembly - Fitch Elmer No. 2230117. One for each cylinder (item 8, 9, 10 or equivalent).
13. Flowmeter, No. 100, Div. of Emerson Electric Company, St. Louis, Mo. (To be used on hydrogen, air and carrier gas lines and to provide flow control to the F.I.D.).
14. Reducator (No. 7221T).
15. Pressure gauge, No. 51N-21, optional.
16. Pressure gauge, No. 9-S11-118, optional.
17. Swagelok or equivalent connectors - Swagelok or equivalent.
18. Swagelok or equivalent fittings - Swagelok or equivalent.
19. Swagelok or equivalent tubing - Swagelok or equivalent.
20. Swagelok or equivalent valves - Swagelok or equivalent.
21. Swagelok or equivalent manifolds - Swagelok or equivalent.
22. Swagelok or equivalent adapters - Swagelok or equivalent.
23. Swagelok or equivalent elbows - Swagelok or equivalent.
24. Swagelok or equivalent tees - Swagelok or equivalent.
25. Swagelok or equivalent unions - Swagelok or equivalent.
26. Swagelok or equivalent elbows - Swagelok or equivalent.
27. Swagelok or equivalent tees - Swagelok or equivalent.
28. Swagelok or equivalent unions - Swagelok or equivalent.
29. Swagelok or equivalent elbows - Swagelok or equivalent.
30. Swagelok or equivalent tees - Swagelok or equivalent.
31. Swagelok or equivalent unions - Swagelok or equivalent.
32. Swagelok or equivalent elbows - Swagelok or equivalent.
33. Swagelok or equivalent tees - Swagelok or equivalent.
34. Swagelok or equivalent unions - Swagelok or equivalent.
35. Swagelok or equivalent elbows - Swagelok or equivalent.
36. Swagelok or equivalent tees - Swagelok or equivalent.
37. Swagelok or equivalent unions - Swagelok or equivalent.
38. Swagelok or equivalent elbows - Swagelok or equivalent.
39. Swagelok or equivalent tees - Swagelok or equivalent.
40. Swagelok or equivalent unions - Swagelok or equivalent.
41. Swagelok or equivalent elbows - Swagelok or equivalent.
42. Swagelok or equivalent tees - Swagelok or equivalent.
43. Swagelok or equivalent unions - Swagelok or equivalent.
44. Swagelok or equivalent elbows - Swagelok or equivalent.
45. Swagelok or equivalent tees - Swagelok or equivalent.
46. Swagelok or equivalent unions - Swagelok or equivalent.
47. Swagelok or equivalent elbows - Swagelok or equivalent.
48. Swagelok or equivalent tees - Swagelok or equivalent.
49. Swagelok or equivalent unions - Swagelok or equivalent.
50. Swagelok or equivalent elbows - Swagelok or equivalent.
51. Swagelok or equivalent tees - Swagelok or equivalent.
52. Swagelok or equivalent unions - Swagelok or equivalent.
53. Swagelok or equivalent elbows - Swagelok or equivalent.
54. Swagelok or equivalent tees - Swagelok or equivalent.
55. Swagelok or equivalent unions - Swagelok or equivalent.
56. Swagelok or equivalent elbows - Swagelok or equivalent.
57. Swagelok or equivalent tees - Swagelok or equivalent.
58. Swagelok or equivalent unions - Swagelok or equivalent.
59. Swagelok or equivalent elbows - Swagelok or equivalent.
60. Swagelok or equivalent tees - Swagelok or equivalent.
61. Swagelok or equivalent unions - Swagelok or equivalent.
62. Swagelok or equivalent elbows - Swagelok or equivalent.
63. Swagelok or equivalent tees - Swagelok or equivalent.
64. Swagelok or equivalent unions - Swagelok or equivalent.
65. Swagelok or equivalent elbows - Swagelok or equivalent.
66. Swagelok or equivalent tees - Swagelok or equivalent.
67. Swagelok or equivalent unions - Swagelok or equivalent.
68. Swagelok or equivalent elbows - Swagelok or equivalent.
69. Swagelok or equivalent tees - Swagelok or equivalent.
70. Swagelok or equivalent unions - Swagelok or equivalent.
71. Swagelok or equivalent elbows - Swagelok or equivalent.
72. Swagelok or equivalent tees - Swagelok or equivalent.
73. Swagelok or equivalent unions - Swagelok or equivalent.
74. Swagelok or equivalent elbows - Swagelok or equivalent.
75. Swagelok or equivalent tees - Swagelok or equivalent.
76. Swagelok or equivalent unions - Swagelok or equivalent.
77. Swagelok or equivalent elbows - Swagelok or equivalent.
78. Swagelok or equivalent tees - Swagelok or equivalent.
79. Swagelok or equivalent unions - Swagelok or equivalent.
80. Swagelok or equivalent elbows - Swagelok or equivalent.
81. Swagelok or equivalent tees - Swagelok or equivalent.
82. Swagelok or equivalent unions - Swagelok or equivalent.
83. Swagelok or equivalent elbows - Swagelok or equivalent.
84. Swagelok or equivalent tees - Swagelok or equivalent.
85. Swagelok or equivalent unions - Swagelok or equivalent.
86. Swagelok or equivalent elbows - Swagelok or equivalent.
87. Swagelok or equivalent tees - Swagelok or equivalent.
88. Swagelok or equivalent unions - Swagelok or equivalent.
89. Swagelok or equivalent elbows - Swagelok or equivalent.
90. Swagelok or equivalent tees - Swagelok or equivalent.
91. Swagelok or equivalent unions - Swagelok or equivalent.
92. Swagelok or equivalent elbows - Swagelok or equivalent.
93. Swagelok or equivalent tees - Swagelok or equivalent.
94. Swagelok or equivalent unions - Swagelok or equivalent.
95. Swagelok or equivalent elbows - Swagelok or equivalent.
96. Swagelok or equivalent tees - Swagelok or equivalent.
97. Swagelok or equivalent unions - Swagelok or equivalent.
98. Swagelok or equivalent elbows - Swagelok or equivalent.
99. Swagelok or equivalent tees - Swagelok or equivalent.
100. Swagelok or equivalent unions - Swagelok or equivalent.

VII. OPERATION OF THE CHROMATOGRAPH (cont'd.)

4. Both solenoid valves are to be interfaced with the timer (or automatic sequencing of the backflush cycle. The timer cam(s) are to be selected and adjusted so that the VCM has all eluted into the second column immediately before the timer initiates backflushing of the first column. This should take place at about 70% of the VCM retention time in minutes. It can be determined by cutting the cycle progressively shorter until a noticeable response drop is experienced on known VCM standards.
5. Install the filter drier assemblies in each cylinder gas line and adjust the gas flow rates to optimize conditions for the chromatographic system being used. The following are typical flow rates used for the Bendix Series 2300 gas chromatograph:
 - a) Nitrogen Carrier Gas - Set the regulator on the cylinder to 50 psig. Adjust the pressure regulator in the N_2 carrier gas line to obtain a VCM chromatograph peak characterized by clean separation from its neighbors. VCM retention time may vary within 2.5 ± 1.0 min. dependent on the particular column system. See also Step 3 above for adjustment of the backflush N_2 flow.
 - b) Detector Air Supply - Set regulator on the cylinder to 80 psig. A Brooks flow controller is to be installed in this line to provide the optimum constant air flow to the detector between 250 to 350 cc/min.
 - c) Hydrogen Supply - Set the cylinder regulator to 20 psig. A Brooks flow controller is to be installed in this line to provide a constant H_2 flow to the detector between 30 and 40 cc/min.
6. Optimum System Temperatures. The following conditions may be used as a starting point:

Column (chromatographic column)	100°C
Injector (heating block)	250°C
Detector (Flame Ionization Detector (FID))	120°C
.....	250°C

When the system is set up, the H_2 flow rate should be adjusted so that the H_2 flow is stable within five seconds. It may be improved by increasing the cylinder pressure. The H_2 flow rate should be adjusted to the desired cylinder pressure.

When the system is set up, the H_2 flow rate should be adjusted so that the H_2 flow is stable within five seconds. It may be improved by increasing the cylinder pressure. The H_2 flow rate should be adjusted to the desired cylinder pressure.

When the system is set up, the H_2 flow rate should be adjusted so that the H_2 flow is stable within five seconds. It may be improved by increasing the cylinder pressure. The H_2 flow rate should be adjusted to the desired cylinder pressure.

VIII. CALIBRATION

Calibration should be performed at least once during each operating period when gas flow rates are changed. It may also be performed at intervals of time. The gas flow rates should be adjusted to the desired values. A standard description of the calibration procedure is given in the following section.

VIII. CALIBRATION (cont'd.)

1. Preparation of Calibration Standards (use fume hood)

Prepare three standard calibration tubes, one each of the 10, 100, and 1000 ppm standard VCN gases, in the following manner:

- a.) Open the standard cylinder and set the outlet regulator at 5 lbe. and open the needle valve at the regulator.
- b.) Set the flow rate, by adjusting the rotameter valve to between 10 and 20 seconds per 10 cc's using the soap film flow meter.
- c.) After the flow rate has been adjusted, connect a pre-conditioned collection tube with the appropriate end toward the cylinder and simultaneously start a stopwatch. NOTE: Do not use Tygon connections between the standard gas cylinder and collection tubes. Substantive steel connections are recommended.

Connect the bubble flow meter to the exit end of the collection tube. While it is being directed to the standard end, using another stopwatch, measure the time in seconds for 100 cc's. Repeat this test three times and record the average time to the nearest 0.1 second.

Remove the calibration tube to the standard for the next test. Repeat this test for the 100 and 1000 ppm standard gases.

Record the results of the calibration test in the calibration data sheet.

Use the following formula to calculate the concentration of the standard gas in the calibration tube:

Concentration (ppm) = (1000 / (Time in seconds / 100)) * (Time in seconds / 100)

Use the three calibration exposures and record the calibration results (see also Note 2).

The calibration tubes in turn to be analyzed through the gas chromatograph analytical system.

After the calibration tubes have been analyzed on a standard gas chromatograph, the calibration of the analytical system is complete.

The calibration of the analytical system is complete.

VIII. CALIBRATION (cont'd.)

3. Preparation of Calibration Curve (cont'd.)

- c.) After analysing and plotting two sets of calibration standards, the best calibration curve can be drawn through the plotted points. It is recommended that the calibration data be analysed using the two variable linear regression capabilities of a Monroe Model 1860 statistical programmable calculator to obtain the equation of the best fit line for this data, and that this line be plotted as a calibration curve. Monroe program No. 114 can be used to generate the equation for the calibration curve.
- d.) Examine the plot of the calibration data carefully. If there appears to be any non-linearity in the log-log plot, it is recommended the curve be segmented into shorter sections, which approach linearity, and that appropriate sample calculation methods are incorporated which are dependent upon VCM level (see Note D).
- e.) It is recommended that subsequent calibration checks be control charts. In this manner significant shifts in calibration which may occur will be readily identified. Following control chart rules will also provide a consistent basis for determining if recalibration is necessary. Refer to the B.F. Goodrich Chemical Company "Control Chart Manual" by R.L. Bowles, A.M. Fairlie, W.D. Hunter and R.S. Murphree, March, 1962. If the daily calibration data violate control chart rules, the analysis system is to be recalibrated and recalibrated until the discrepancy is detected and recalibration is necessary.

The calibration curve used in this analysis is the one prepared by J.M. Whitney and F.V. Zemanek, and is the one used in the Government Monitoring VCM Reports from the VCM Laboratory. A copy of this report is obtained from the VCM Laboratory or the N.S. Laboratory.

The preparation of calibration standards which cover the range of VCM levels and provide excellent reproducibility. This is done by using the certified calibration gas and varying flow rate to obtain different calibration levels. The calibration gas is used to obtain the calibration curve that will be used in the analysis. In this report, a standard VCM calibration curve is shown for the VCM Laboratory.

The calibration curve is used to determine the VCM level of the sample. The calibration curve is used to determine the VCM level of the sample. The calibration curve is used to determine the VCM level of the sample. The calibration curve is used to determine the VCM level of the sample.

IX. SAMPLE ANALYSIS PROCEDURE

3. The collection tube is allowed to desorb for 2 minutes, during which time the recorder baseline, attenuation, and/or integration parameters can be appropriately checked and adjusted.
4. After the desorption step is completed the sample is injected onto the analytical G.C. column and the backflush timer switch is set to "reset". The recorder and/or integrator is also started simultaneously.
5. If using a recorder only, attenuate as necessary to keep the peak height of all peaks in the upper half of the chart. Record the sample identification, analysis start, and attenuation of each peak on the recorder chart. If using an integrator, the logarithmic scale attenuator can be used to keep all peaks on the recorder chart. After 30 seconds have elapsed, switch the timer to "Auto". This will return the backflush timer to the start of the cycle at the completion of the analysis.
6. The recorder will graph the response with time of each low boiling component in the collection tube being analyzed. The backflush of the first analytical column will be automatically actuated by the preset timer immediately after the VCM has eluted from the first column.
If using the Century system, repeat injections may be made on aliquots of single desorbed collection tubes.
7. Record the sample identification, collection tube I.D. number, attenuation, peak width and height, (or area counts if an integrator is used) and the sample desorb volume (supplied with sample) on the appropriate laboratory data collection form (see Attachment #3).

X. CALCULATION OF RESULTS (cont'd.)

The above equation can be easily calculated using Monroe Program No. 113.1840.

NOTE 1: Although sampling methods are not outlined in this test procedure, it is understood that the capacity of the collection tube has not been exceeded.

It has been found that collection periods characterized by high air flow rates (5000 cfm.), and/or high humidity may lower the capacity of the collection tube dependent upon the length of the collection period and the concentration level being sampled.

NOTE 2: The results should be appropriately adjusted relative to the ambient barometric coefficients. See also VIII, 2.

0.1000

Attachment No. 1
Physical Collection Index, Internal Description
Sample Calibration Curve
US phase log-log

-88-
Attachment #2
Monroe Program No. 1121860 Perseant

AAAA	001	1	0030	060	
	002	2	1	024	+
	003		2	046	
	004		3	040	
	005		4	074	+
	006		5	002	+
	007		6	010	+
	008		7	034	+
	009		8	034	+
	010		9	034	+
	011		0	034	+
	012		1	034	+
	013		2	034	+
	014		3	034	+
	015		4	034	+
	016		5	034	+
	017		6	034	+
	018		7	034	+
	019		8	034	+
	020		9	034	+

⁻⁶²
Attachment #3
Marine Program No. 13,1560 Patent

1990	1991	1
1991	1992	2
1992	1993	3
1993	1994	4
1994	1995	5
1995	1996	6
1996	1997	7
1997	1998	8
1998	1999	9
1999	2000	10

Attachment 14
Monica Program No. 14/140 Printed

2000	001	1	2000	067
1	001	1	2	099
2	004	0	3	120
3	067		4	001
4	176		5	001
5	042	A	6	066
6	114		7	042
7	201		8	050
8	187		9	120
9	072		10	101
10	200		11	027
11	000		12	000
12	000		13	000
13	000		14	000
14	000		15	000
15	000		16	000
16	000		17	000
17	000		18	000
18	000		19	000
19	000		20	000
20	000		21	000
21	000		22	000
22	000		23	000
23	000		24	000
24	000		25	000
25	000		26	000
26	000		27	000
27	000		28	000
28	000		29	000
29	000		30	000
30	000		31	000
31	000		32	000
32	000		33	000
33	000		34	000
34	000		35	000
35	000		36	000
36	000		37	000
37	000		38	000
38	000		39	000
39	000		40	000
40	000		41	000
41	000		42	000
42	000		43	000
43	000		44	000
44	000		45	000
45	000		46	000
46	000		47	000
47	000		48	000
48	000		49	000
49	000		50	000
50	000		51	000
51	000		52	000
52	000		53	000
53	000		54	000
54	000		55	000
55	000		56	000
56	000		57	000
57	000		58	000
58	000		59	000
59	000		60	000
60	000		61	000
61	000		62	000
62	000		63	000
63	000		64	000
64	000		65	000
65	000		66	000
66	000		67	000
67	000		68	000
68	000		69	000
69	000		70	000
70	000		71	000
71	000		72	000
72	000		73	000
73	000		74	000
74	000		75	000
75	000		76	000
76	000		77	000
77	000		78	000
78	000		79	000
79	000		80	000
80	000		81	000
81	000		82	000
82	000		83	000
83	000		84	000
84	000		85	000
85	000		86	000
86	000		87	000
87	000		88	000
88	000		89	000
89	000		90	000
90	000		91	000
91	000		92	000
92	000		93	000
93	000		94	000
94	000		95	000
95	000		96	000
96	000		97	000
97	000		98	000
98	000		99	000
99	000		100	000

Memo Program No 114.160 Pimentout (Cont'd)

0100	001	1
1	001	2
2	011	3
3	121	4
4	001	5
5	002	6
6	020	7
7	121	8
8	001	9
9	004	10
10	011	11
11	001	12
12	001	13
13	002	14
14	020	15
15	121	16
16	001	17
17	001	18
18	001	19
19	001	20
20	001	21
21	001	22
22	001	23
23	001	24
24	001	25
25	001	26
26	001	27
27	001	28
28	001	29
29	001	30
30	001	31
31	001	32
32	001	33
33	001	34
34	001	35
35	001	36
36	001	37
37	001	38
38	001	39
39	001	40
40	001	41
41	001	42
42	001	43
43	001	44
44	001	45
45	001	46
46	001	47
47	001	48
48	001	49
49	001	50
50	001	51
51	001	52
52	001	53
53	001	54
54	001	55
55	001	56
56	001	57
57	001	58
58	001	59
59	001	60
60	001	61
61	001	62
62	001	63
63	001	64
64	001	65
65	001	66
66	001	67
67	001	68
68	001	69
69	001	70
70	001	71
71	001	72
72	001	73
73	001	74
74	001	75
75	001	76
76	001	77
77	001	78
78	001	79
79	001	80
80	001	81
81	001	82
82	001	83
83	001	84
84	001	85
85	001	86
86	001	87
87	001	88
88	001	89
89	001	90
90	001	91
91	001	92
92	001	93
93	001	94
94	001	95
95	001	96
96	001	97
97	001	98
98	001	99
99	001	100

NAME _____ DOB _____
ADDRESS, PHONE, EMAIL _____

NAME _____ DOB _____
Address _____
E-mail _____

	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7
	1	2	3	4	5	6	7

000000

-73-
VINYL CHLORIDE IN AIR

Physical and Chemical Analysis Branch

Analytical Method

Analyte: Vinyl Chloride

Method No.: PACAM 178

Matrix: Air

Range: 0.008 to 5.2 mg/m³ in
a 5-liter air sample

Procedure: Adsorption on
activated carbon, desorption
with carbon disulfide, gas
chromatography

Precision (CV%): 0.08 at
levels of 7 and 71 mg/m³

Date issued: 9/3/74

Classification: B (accepted)

Date revised: 1/29/76

1. Principle of the Method

A known volume of air is drawn through two small activated carbon tubes containing activated carbon beads. Vinyl chloride present in the air is adsorbed on the activated carbon beads. The activated carbon beads are then desorbed with carbon disulfide, and the resulting solution is injected into a gas chromatograph equipped with a flame ionization detector. The detector response is proportional to the amount of vinyl chloride present in the original air sample.

3. Interferences

- 3.1 When the amount of water in the air is so great that condensation actually occurs in the tube, organic vapors will not be trapped effectively. Experiments indicate that high humidity severely decreases the capacity of activated carbon for organic vapors.
- 3.2 When two or more substances are known or suspected to be present in the air, such information, including their suspected identities, should be transmitted with the sample since these compounds may interfere with the analysis for vinyl chloride.
- 3.3 Any compound that has the same retention time as vinyl chloride at the operating conditions described in this method is an interference. Hence, retention time data on a single column, or even on a number of columns, may not provide proof of chemical identity. Often, operating conditions can be modified to eliminate interferences. Samples should be analyzed by an independent method when overlapping gas chromatographic peaks cannot be resolved.

4. Precision and Accuracy

- 4.1 The precision of retention timing from the analysis of two lots of sorbent tubes, one lot of 20 tubes exposed to a vinyl chloride concentration of 7.3 mg/m^3 in air and another lot of 20 tubes exposed to a concentration of 71.3 mg/m^3 , were 6.07% and 1.07%, respectively. These values reflect total sampling and analytical error as well as retention timing precision errors.

4.2 The accuracy of retention timing is given some indication of the accuracy, although not the precision, of the analysis of a primary standard. These standards were analyzed in the same manner as the samples and were found to be within 1.0% of the expected value. The accuracy of the analysis of the standards was determined by the use of a gravimetric method.

Page 10 of 10

-5-

6. Apparatus

6.1 **Personal Sampling Pump.** The pump should be a properly calibrated personal sampling pump for personal and area samples. It should be calibrated with a representative sorbent tube in the sampling line. A dry or wet test meter or a glass rotameter that will determine the flow rate (50 ml/min) to within $\pm 5\%$ may be used for the calibration.

6.2 **Sorbent Tubes.** The glass tubes have both ends flame sealed. Each is 7 cm long, 6-mm o.d., 4-mm i.d., and contains two sections of 20/40-mesh activated carbon separated by a 2-mm portion of sorbent foam. The activated carbon is prepared from coconut shells and is fired at 600°C prior to packing to remove adsorbed materials. The primary absorbing section contains 100 mg of sorbent, the backup section 50 mg. A 2-mm portion of sorbent foam is placed between the outlet end of the tube and the backup section. A plug of silanized glass wool is placed in front of the absorbing section. The pumping device across the tube must be less than 2 in. of water at a flow rate of 0.2 l/min.

6.3 **Gas Chromatograph equipped with a flame ionization detector.**

The gas chromatograph should be equipped with a flame ionization detector. The detector should be calibrated with a known concentration of the gas to be measured. The gas flow rates should be adjusted to give the best possible resolution and sensitivity.

The gas chromatograph should be equipped with a flame ionization detector. The detector should be calibrated with a known concentration of the gas to be measured. The gas flow rates should be adjusted to give the best possible resolution and sensitivity.

6.4 **Flow rate of 0.2 l/min (type graduated to 0.1-ml increments).**

The flow rate of 0.2 l/min should be maintained throughout the sampling process. It is preferable to

1.5 Regulated hydrogen.

1.6 Filtered compressed air.

2. Procedure

2.1 Cleaning of Equipment. All glassware used for the laboratory analysis should be washed with detergent and thoroughly rinsed with distilled water.

2.2 Collection and Shipping of Samples

2.2.1 Immediately before sampling, the ends of two tubes are broken to provide an opening of least one-half the internal diameter of the tube (2 mm).

2.2.2 The second shortest tube is used as a backup and is positioned next to the sampling pump in tandem with the first tube.

2.2.3 The tubes are placed in a vertical position with the impregnation of reagent material in the tubes being so positioned as to minimize absorption of the gas sample during the process.

2.2.4 The tubes are not to be placed through any hose or other tubing which may cause leakage.

2.2.5 The tubes are not to be placed in a position where they are exposed to direct sunlight or other heat source, and to prevent as completely as possible the tubes from being at a flow rate of 20 ml/min. The maximum flow rate for the tubes is 5 ml/min. (See Section 2.1.)

Capped tubes are packed tightly before they are shipped to the laboratory. The use of two tubes in series has eliminated the need for cooling during shipping. However, if two tubes are not used, i.e., only one tube is used, and if the samples will spend a day or more in transit, then cooling (i.e., with Dry Ice) is necessary to minimize migration of vinyl chloride to the backup section.

Samples received at the laboratory are logged in and immediately stored in a freezer (around -20°C) until time for analysis. Samples may be stored in this manner for long periods of time with no appreciable loss of vinyl chloride (2 months). Vinyl chloride will equilibrate between the two sections of backup system, i.e., it will migrate to the backup section. The migration is dependent upon time, i.e., it will migrate in 2 weeks and may be confined with samples

3211

3210

113

Injection. The first step in the analysis is the injection of the sample into the gas chromatograph. To eliminate difficulties arising from solvent back or distillation within the syringe needle, one should employ the solvent flush injection technique. The 10- μ l syringe is first flushed with solvent several times to wet the barrel and plunger. Two microliters of solvent is drawn into the syringe to increase the accuracy and reproducibility of the injected sample volume. The needle is removed from the solvent and the plunger is pulled back about 0.4 μ l to separate the solvent flush from the sample with a pocket of air to be used as a marker. The needle is then immersed in the sample, and a 3- μ l aliquot is withdrawn to the 7.4 μ l mark (2 μ l solvent + 0.4 μ l air + 5 μ l sample = 7.4 μ l). After the needle is removed from the sample and prior to injection the plunger is pulled back a short distance to minimize evaporation of the sample from the tip of the needle. Duplicate injections of each sample and standard are made. No more than a 3% difference in area is to be expected. Automatic sampling devices may also be used.

234

Measurement of Area. The area under the sample peak is measured by an electronic integrator or some other suitable form of area measurement, and preliminary results are read from a standard curve prepared as discussed below.

THE UNIVERSITY OF CHICAGO

the form and quantity of matter and also from one body to another. The 2nd principle is promulgated at least once in the Bible. The 3rd principle is the Creation myth. The 4th principle is the story of the fall of man. The 5th principle is the story of the flood. The 6th principle is the story of the tower of Babel. The 7th principle is the story of the exodus. The 8th principle is the story of the conquest of Canaan. The 9th principle is the story of the judges. The 10th principle is the story of the kings. The 11th principle is the story of the prophets. The 12th principle is the story of the messiah. The 13th principle is the story of the resurrection. The 14th principle is the story of the ascension. The 15th principle is the story of the Pentecost. The 16th principle is the story of the church. The 17th principle is the story of the end times. The 18th principle is the story of the new heaven and new earth. The 19th principle is the story of the new Jerusalem. The 20th principle is the story of the new creation.

1. Gas Sampling Efficiency. Strikant takes from the gas stream a known volume of gas. Weighing scales are used in this determination. The known volume of gas is injected into a bag containing a known volume of air. The bag is made of Tuller (or a material that will not react chemically and not absorb it) and should have a gas sampling inlet and a release-inflating port. The concentration in the bag may be found if mass, volume, and total pressure are known. A measured volume of air is drawn through the bag with a calibrated sampling pump. An inlet valve on the pump is closed. Then the valve on the bag is closed and the air is released. The gas is then injected into the gas stream. The concentration in the bag is compared to the concentration in the stream.

The desorption efficiency equals the amount of vinyl chloride desorbed from the solvent divided by the quantity of vinyl chloride contained in the volume of synthetic atmosphere sampled, or

$$\frac{(\text{amount of vinyl chloride from solvent})}{(\text{volume of vinyl chloride in bag}) \times (\text{vol of atmosphere sampled})}$$

9. Calibration and Standards

CAUTION: Laboratory Operations Involving Carcinogens

Vinyl chloride has been identified as a human carcinogen and appropriate precautions must be taken in handling this gas. The Occupational Safety and Health Administration has promulgated regulations for the use and handling of vinyl chloride. This may be found in 29 CFR 1910.97a (Section 1910.97a is Title 29 of the Code of Federal Regulations available in the Federal Register, Vol. 39, No. 234, dated October 4, 1974, pp. 52000-52002).

Vinyl chloride is a colorless gas with a sweet odor. It is highly flammable and explosive. It is also a known carcinogen. Therefore, appropriate precautions must be taken in handling this gas. The Occupational Safety and Health Administration has promulgated regulations for the use and handling of vinyl chloride. This may be found in 29 CFR 1910.97a (Section 1910.97a is Title 29 of the Code of Federal Regulations available in the Federal Register, Vol. 39, No. 234, dated October 4, 1974, pp. 52000-52002).

Vinyl chloride is a colorless gas with a sweet odor. It is highly flammable and explosive. It is also a known carcinogen. Therefore, appropriate precautions must be taken in handling this gas. The Occupational Safety and Health Administration has promulgated regulations for the use and handling of vinyl chloride. This may be found in 29 CFR 1910.97a (Section 1910.97a is Title 29 of the Code of Federal Regulations available in the Federal Register, Vol. 39, No. 234, dated October 4, 1974, pp. 52000-52002).

2.1.3

Volumetric Method. A 1-ml gas sample of pure vinyl chloride is drawn into a gas-tight syringe and the valve is closed. The tip of the needle is inserted into a 10-ml volumetric flask containing approximately 5 ml of CS_2 . The valve is opened and the plunger is withdrawn slightly to allow the CS_2 to enter the syringe. The action of the vinyl chloride dissolving in the CS_2 creates a vacuum and the syringe becomes filled with the solvent. An air bubble (2%) is present and has been found to be due to the void volume in the needle of the syringe. The solution is returned to the flask and the syringe is rinsed with clean CS_2 and the washings added to the volumetric flask. The volumetric flask is then filled to the mark with CS_2 . Other standards are then prepared from this stock solution.

Standards are stored in a freezer at $-20^\circ C$ and have been found to be stable at this temperature for 3 days. Tight-fitting plastic caps on the volumetric flask must be used to retain the vinyl chloride better than ground-glass stoppers.

10.3 The concentration of vinyl chloride in air may be expressed in mg/m^3 :

$$\text{mg/m}^3 = \frac{\text{corrected volume (in ml)}}{\text{volume of air sampled (in l)}}$$

10.4 The concentration may also be expressed in terms of parts per million (ppm) by using:

$$\text{ppm} = \text{mg/m}^3 \times \frac{24.45}{\text{M.W.}} \times \frac{760}{P} \times \frac{1.01325}{298}$$

where: 24.45 = molar volume (l/mole) at 25°C and 760 mmHg.

M.W. = molecular weight.

P = pressure (mmHg) of air sampled.

T = temperature (°K) of air sampled.

Prepared by: R. W. Tarr, and W. J. Woodfin.

Environmental Engineering
Department of Health, Education and Welfare
Washington, D. C. 20460

-84-

Mr. Ray Neval
Klineclander Paper Company

-2-

August 24 1980

The results above are based on limited testing and actual continuing factory exposure may be different. The test results are believed to be representative of conditions found under the particular and peculiar operating conditions existing at the time and place where the tests were made and no representations are made concerning other times, conditions or production levels. Submission of the test results is not intended to determine or insure compliance with any standard, specification, governmental regulation or policy, although such results can be useful in establishing the elements necessary for a program of compliance. We hope that the data presented will assist you in continuing to provide a safe workplace for your employees.

The responsibility of maintaining an environment in compliance with OSHA regulations rests with the individual manufacturer.

Sincerely yours,

AMERICAN CHEMICAL GROUP
Labor Division

W. B. Setava
Product Manager

20000003

CONFIDENTIAL

On August 7, 1980 the Rhineland Paper Company, Rhineland, Wisconsin was visited to monitor the workplace for VCM. The visit was made at the request of the customer, since they were using a new latex in their operations. Samples were collected on six charcoal tubes at 30-60 cc/min. Analysis was by gas chromatography.

Samples were collected from the breathing zone of all employees working in the latex operation. General area samples were collected in the vicinity of the release. Analysis of the samples was done by the Environmental Health Laboratory in Madison, Wisconsin.

Analysis of the samples was done by the Environmental Health Laboratory in Madison, Wisconsin. The laboratory is located at 1000 University Avenue, Madison, Wisconsin 53706. The laboratory is a part of the University of Wisconsin-Madison.

RHINELAND PAPER CO. - VCM

T. S. Stahl

AXTON - D/0020, 5-H

8-25-80

Cleveland Chemical

K. Schneider

REINELANDER PAPER COMPANY
PCM MONITORING RESULTS - 8/7/80

<u>Sample/Location/ Operation</u>	<u>Sample Time (min.)</u>	<u>Sample Volume (liters)</u>	<u>Concentration (ppm)</u>
H. Fisher	200	10.2	0.2
Coating/Mixing	223	10.2	0.3
H. Craig, Foreman	426	10.1	0.2
J. ...	130	10.0	0.1
...	220	11.3	0.1
...	170	8.0	0.3
...	...	...	0.2

**SAFETY AND HEALTH
SERVICES PROGRAM**

Wausau Insurance Companies

2004 ACCTODOO GAME
WAUSAU, WISCONSIN 54401

C. Young, Vice President

SECRET

Company, Incorporated

CONFIDENTIAL

SECRET

315 West Springfield Street
Burlington, Vermont

August 18, 1960

ALL INFORMATION CONTAINED HEREIN IS UNCLASSIFIED; DATE 08-29-2007 BY 60322 UCBAW/SJS

-44-
-3-
U. Thomas Myers
Chiceland Paper Company
Incorporated

August 18, 1968

time-weighted average permissible exposure level of one part per million
(p.p.m.) and also below the OSHA action level of 0.5 p.p.m.

-4-

2004.1

RECEIVED AT 1700. 20.000

RECEIVED AT 1700. 20.000
RECEIVED AT 1700. 20.000
RECEIVED AT 1700. 20.000
RECEIVED AT 1700. 20.000

RECEIVED AT 1700. 20.000
RECEIVED AT 1700. 20.000
RECEIVED AT 1700. 20.000
RECEIVED AT 1700. 20.000

Lower Limit of
Resistance (2)
kPa (3)

Sample Time

7:25 a.m.-0:17 a.m.
7:25 a.m.-0:17 a.m.
7:43 a.m.-0:23 a.m.
8:16 a.m.-0:31 a.m.
8:23 a.m.-0:40 a.m.
8:33 a.m.-0:47 a.m.
8:38 a.m.-0:46 a.m.
9:17 a.m.-0:50 a.m.
9:16 a.m.-0:50 a.m.
10:01 a.m.-10:26 a.m.

0.2
0.2
0.2
0.3
0.2
0.3
0.2
0.2
0.3
0.3

1
1
1
1
1
1
1
1
1
1

TABLE I (Continued)

- 4 -

Sample No.	Sample Location	Concentration of Vapor Chloride		Lower Limit of Detection (3) Ref. (1)	Sample Time
		Found	Ref. (2)		
77	Jack Anderson	Detected	1	0.2	10:15 a.m.-10:52 a.m.
78	Jack Anderson	Detected	1	0.2	10:35 a.m.-11:10 a.m.
79	James Flaherty	Detected	2	0.2	12:20 p.m.-1:15 p.m.
80	James Flaherty	Detected	2	0.2	2:15 p.m.-2:46 p.m.
81	Jack Anderson	Detected	2	0.4	2:16 p.m.-2:40 p.m.

Notes - Parts of samples analyzed by other means of air by volume.

also detectable by means of the 20 min. of Federal Regulations, Part 1910, Subpart Z, Section 1910.1007, vapor chloride.

Lower limit of detection in parts per million, analytical method and volume of air sampled.

45828