

VC STACK MONITOR AUTOMATIC BENCHMARKING

CHECK	ACTION STEPS	COMMENTS (below steps)
-------	--------------	------------------------

THE VC ANALYZER CALIBRATES AND RECORDS HOURLY VC CONCENTRATIONS ON BOTH INCINERATOR STACKS. REFER TO SOP76 FOR HIGH VC HOURLY CONCENTRATIONS.

THE AUTOMATIC BENCHMARKING (CALIBRATION) OF THE VC ANALYZER IS PROGRAMED TO BE DONE REMOTELY FROM TELI EVERY DAY TO CHECK THE INSTRUMENT FOR ACCURACY.

- 1. AT APPROXIMATELY 8:15 EVERY DAY, THE ANALYZER WILL AUTOMATICALLY BENCHMARK ITSELF FOR ACCURACY.
- 2. AFTER THREE BENCHMARKING RUNS, THE ANALYZER WILL RESUME NORMAL SAMPLING.
- 3. AFTER THE BENCHMARKING RUNS, THE OPERATOR WILL RECORD THE DEVIATION NUMBERS ON THE LOGSHEET.
- 4. IF THE DEVIATION IS GREATER THAN OR LESS THAN 5%, THE LEAD OPERATOR IS CALL A MEMBER OF THE ANALYZER GROUP TO CHECK THE SYSTEM. IF IT IS ON THE WEEKENDS OR HOLIDAY, THE LEAD OPERATOR IS TO NOTIFY THE SHIFT SUPERVISOR TO CALL OUT AN ANALYZER REPAIRMAN TO CORRECT THE CALIBRATION DEVIATION.

CONFIDENTIAL:
Subject to Protective Order
of 14th Judicial District Court
No. 91-1145

SL 001694