

Name: Mc Allister
Next test due:

TSI MODEL 8015 PORTACOUNT FIT-TESTING SOFTWARE
Version 2.01 S/N 927
VISTA POLYMERS

FIT TEST REPORT

Test date: 08-03-1994
Test time: 7:17:32 am
Test subject last name: Mc Allister
Test subject first name: Billy
ID number: 427722276
Operator name: H Pruitt
Next fit test due:
Test agent: Ambient particles
Test device: TSI MODEL 8010 PORTACOUNT

RESPIRATOR ID

Size: Small
Model: Ultravue Facepiece
Manufacturer: MSA
Approval number: TC-19C-113

Additional notes:
Additional notes:

TEST DATA

Pass/Fail level: 100
Test cycles.....: 1

Ex.	Ambient (Part/cc)	Mask (Part/cc)	Fit Factor	Pass/Fail
NB	22000	0.310	71100	PASS
DB	24650	1.230	20000	PASS
SS	25600	0.290	88400	PASS
UD	25450	0.540	47200	PASS
R	26400	1.750	15100	PASS
NB	25950	0.360	72100	PASS

OVERALL FIT FACTOR = 33964 PASS
(Calculated per ANSI Z88.2-1980)

Operator Helene Pruitt Date 8-3-94

Test Subject Billy McAllister Date 8-3-94