

DAILY STATUS REPORT
GAS CHROMATOGRAPH

DATE 8 AM 12/17/78 TO 8 AM 12/18/78

BLDG. 121	8-4 SHIFT		4-12 SHIFT		12-8 SHIFT	
	MONITOR <u>Beckham</u>	TIME	MONITOR <u>Cooksey</u>	TIME	MONITOR <u>Munley</u>	TIME
Standard Gas Span Check	<u>—</u> / <u>—</u>	<u>1445</u>	<u>—</u> / <u>—</u>	<u>—</u>	<u>—</u> / <u>—</u>	<u>—</u>
If Out of Range Was Prod. Foreman Notified	<u>YES</u> Foreman	<u>NO</u>	<u>YES</u> Foreman	<u>NO</u>	<u>YES</u> Foreman	<u>NO</u>
Downtime In Minutes If None Write None	<u>51 min.</u>		<u>NONE</u>		<u>NONE</u>	
If GC Is Not Operating Properly, Note Condition	<u>H-2200</u> <u>N-1700</u> <u>ran out of Hydrogen</u>				<u>could not span check</u> <u>auto sampler has</u> <u>been removed</u>	
If OK, Write OK	<u>OK</u>		<u>O.K.</u>		<u>OK</u>	
Recorder						
If Other Than OK Note Condition						
If OK, Write OK	<u>OK</u>		<u>OK</u>		<u>OK</u>	

NOTE ANY OTHER UNUSUAL CONDITIONS ABOUT EQUIPMENT.

LPA

Standard Gas Span Check	TIME		TIME		TIME	
	<u>—</u> / <u>—</u>	<u>1500</u>	<u>—</u> / <u>—</u>	<u>—</u>	<u>3.0</u> / <u>3.47%</u>	<u>0140</u>
If Out of Range Was Prod. Foreman Notified	<u>YES</u> Foreman	<u>NO</u>	<u>YES</u> Foreman	<u>NO</u>	<u>YES</u> Foreman	<u>NO</u>
Downtime In Minutes If None, Write None	<u>NONE</u>		<u>NONE</u>		<u>NONE</u>	
If GC Is Not Operating Properly, Note Condition.	<u>H-75</u> <u>N-1800</u>					
If OK, Write OK.	<u>OK</u>		<u>O.K.</u>		<u>OK</u>	
Recorder						
If Other Than OK Note Condition	<u>OK</u>					
If OK, Write OK			<u>O.K.</u>		<u>OK</u>	

NOTE ANY OTHER UNUSUAL CONDITIONS ABOUT EQUIPMENT.

TANK FARM

Span Check	<u>—</u> / <u>—</u>	<u>—</u> / <u>—</u>	<u>101.0</u> / <u>98.0</u> / <u>0215</u>
Was Foreman Notified	<u>Yes - No</u>	<u>Yes - No</u>	<u>Yes - No</u>
Downtime in Minutes		<u>NONE</u>	<u>NONE</u>
THA Status	<u>H-1550</u>		
If OK, Write OK		<u>O.K.</u>	<u>OK</u>
Recorder Status			
If OK, Write OK		<u>OK</u>	<u>OK</u>