

Name: Holcomb
Next test due:

TSI MODEL 8015 PORTACOUNT FIT-TESTING SOFTWARE
Version 2.01 S/N 927
VISTA POLYMERS

FIT TEST REPORT

Test date: 08-02-1994
Test time: 1:48:44 pm
Test subject last name: Holcomb
Test subject first name: ~~Charles~~ *BILL*
ID number: 426844601
Operator name: S. Carter
Next fit test due:
Test agent: Ambient particles
Test device: TSI MODEL 8010 PORTACOUNT

RESPIRATOR ID

Size: Medium
Model: Ultravue Facepiece
Manufacturer: MSA
Approval number: TC-19C-113

Additional notes:
Additional notes:

TEST DATA

Pass/Fail level: 100
Test cycles.....: 1

Ex.	Ambient (Part/cc)	Mask (Part/cc)	Fit Factor	Pass/Fail
NB	4400	0.240	18300	PASS
DB	4095	0.400	10200	PASS
SS	3865	1.110	3480	PASS
UD	3585	2.590	1380	PASS
R	3320	0.850	3900	PASS
NB	3115	0.100	31100	PASS

OVERALL FIT FACTOR = 4129 PASS
(Calculated per ANSI Z88.2-1980)

Operator *Shelby D. Carter* Date *8-2-94*

Test Subject *William Holcomb* Date *8-2-94*