DIRECTIONS FOR TEST NO. 14-49-45

SUBJECT: ~~SPECIFIC GRAVITY OF AROCLOR~~

METHOD: ~~HYDROMETERS~~

Apparatus: Appropriate hydrometer. C.J. Tagliabue Co.
Cylinder
Thermometer
Constant temperature bath.

Procedure: Have constant temperature bath between 66 -67°C.

Heat Arocolor in 600°C beaker to 70°C. Pour into gravity cylinder to a height just sufficient to float hydrometer. Stir with thermometer until 65.5°C. Place cylinder on constant temperature bath and stir until temperature reaches 65°C. Insert hydrometer and allow it to settle to its own level. Read stem at bottom of meniscus thru water and arocolor. Report to nearest .0005.

DSW 330781

STLCOPCB4077425