

Name: Pace
Next test due:

TSI MODEL 8015 PORTACOUNT FIT-TESTING SOFTWARE
Version 2.01 S/N 927
VISTA POLYMERS

FIT TEST REPORT

Test date: 08-12-1994
Test time: 10:27:34 am
Test subject last name: Pace
Test subject first name: O. C.
ID number: 426766938
Operator name: S. Carter
Next fit test due:
Test agent: Ambient particles
Test device: TSI MODEL 8010 PORTACOUNT

RESPIRATOR ID

Size: Medium
Model: Ultravue Facepiece
Manufacturer: MSA
Approval number: TC-19C-113

Additional notes:
Additional notes:

TEST DATA

Pass/Fail level: 100
Test cycles.....: 1

Ex.	Ambient (Part/cc)	Mask (Part/cc)	Fit Factor	Pass/Fail
NB	41350	0.140	295000	PASS
DB	40550	0.540	75100	PASS
SS	39050	0.560	69800	PASS
UD	37950	0.500	75900	PASS
R	36550	2.360	15500	PASS
NB	35100	0.530	66300	PASS

OVERALL FIT FACTOR = 48463 PASS
(Calculated per ANSI Z88.2-1980)

Operator Shelby D. Carter Date 8-12-94

Test Subject O.C. Pace Date 8-12-94