

Name: HARGROVE
Next test due:

TSI MODEL 8015 PORTACOUNT FIT-TESTING SOFTWARE
Version 2.01 S/N 927
VISTA POLYMERS

FIT TEST REPORT

Test date: 10-12-1994
Test time: 8:55:04 am
Test subject last name: HARGROVE
Test subject first name: TIMOTHY
ID number: 587061283
Operator name: STEWART
Next fit test due:
Test agent: Ambient particles
Test device: TSI MODEL 8010 PORTACOUNT

RESPIRATOR ID

Size: SMALL
Model:
Manufacturer: 3M
Approval number: TC-21C-488

Additional notes:
Additional notes:

TEST DATA

Pass/Fail level: 100
Test cycles.....: 1

Ex.	Ambient (Part/cc)	Mask (Part/cc)	Fit Factor	Pass/Fail
NB	4970	1.330	3740	PASS
DB	4635	0.120	38600	PASS
SS	4315	0.160	26900	PASS
UD	4005	0.700	5730	PASS
R	3645	2.560	1420	PASS
NB	3400	0.500	6800	PASS

OVERALL FIT FACTOR = 4424 PASS
(Calculated per ANSI Z88.2-1980)

Operator Arthur J. Stewart Date 10-12-94

Test Subject Timothy R. Hargrove Date 10-12-94