

Milk Sample

Aroclor 1254

DDT

R.T. min	P.H. mm		R.T. min	P.H. mm		R.T. min	P.H. mm	
2.5	142	X	-3	2.8	21	0.2	2.6	59
3.1	28	X						
3.5	92	X						
4.0	10	X						
4.8	174	X				4.4	22	
5.9	2128	X	+3	5.6	25	0.3	H	
6.8	2928	X	-1	6.9	286	3.2	H	
8.3	416	X	+1	8.2	126	1.4	H	
9.3	80	2.9	+3	9.0	96	1.1	✓ 9.0	83
10.3	100	3.0	+2	10.1	576	6.4	✓ 10.0	79
11.1	66	2.4	-5	11.6	1220	14.2	✓ 11.6	186
12.1	138	5.0						
12.5	9	3.2	+4	13.1	132	1.5	14.2	756 ✓
15.1	37	1.3	+3	14.8	624	6.9		
16.3	59	2.1	+3	16.0	1040	11.6		
18.2	171	6.2	+3	17.9	1336	14.8		
20.3	119	4.3	+4	19.9	860	9.6	20.9	472
22.3	40	1.4	+4	22.8	748	8.3		
24.1	27	1.0						
26.2	126	4.5	+4	25.6	1000	11.1	25.3	616
28.6	15	0.6	+9	27.7	174	1.9		
30.6	12	0.4						
33.7	13	0.4	+8	32.9	240	2.7		
				36.2	186	2.1		
38.0	97	3.5	-5	38.5	157	1.7		
				50.4	88	1.0		

100.0

① data show does not absolutely prove the presence of PCBs in milk sample

② The "less than .6 ppm" is probably on a whole milk basis... might be as low as .06 range.