

Name: Ostrowski

Next test due:

TSI MODEL 8015 PORTACOUNT FIT-TESTING SOFTWARE  
Version 2.01 S/N 927  
VISTA POLYMERS

## FIT TEST REPORT

-----  
Test date: 08-02-1994  
Test time: 1:31:19 pm  
Test subject last name: Ostrowski  
Test subject first name: Gordon  
ID number: 427823371  
Operator name: S. Carter  
Next fit test due:  
Test agent: Ambient particles  
Test device: TSI MODEL 8010 PORTACOUNT

## RESPIRATOR ID

Size: Large  
Model: Ultravue Facepiece  
Manufacturer: MSA  
Approval number: TC-19C-113

Additional notes:  
Additional notes:

## TEST DATA

-----  
Pass/Fail level: 100  
Test cycles.....: 1

| Ex. | Ambient<br>(Part/cc) | Mask<br>(Part/cc) | Fit Factor | Pass/Fail |
|-----|----------------------|-------------------|------------|-----------|
| NB  | 5900                 | 0.190             | 31000      | PASS      |
| DB  | 5865                 | 0.680             | 8630       | PASS      |
| SS  | 5990                 | 0.650             | 9220       | PASS      |
| UD  | 6010                 | 0.440             | 13600      | PASS      |
| R   | 5760                 | 7.080             | 814        | PASS      |
| NB  | 5570                 | 0.730             | 7630       | PASS      |

OVERALL FIT FACTOR = 3551 PASS  
(Calculated per ANSI Z88.2-1980)

Operator Shelby D. Carter Date 8-2-94  
Test Subject Gordon Ostrowski Date 8-2-94